

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091323\
 Data File : BN027619.D
 Acq On : 12 Sep 2023 20:43
 Operator : MA/JU
 Sample : 04287-10
 Misc :
 ALS Vial : 121 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BR-03-127-090123-3V

Quant Time: Sep 13 06:03:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:45:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.037	152	6288	0.400	ng	0.00
7) Naphthalene-d8	10.852	136	16870	0.400	ng	0.00
13) Acenaphthene-d10	14.661	164	7988	0.400	ng	0.00
19) Phenanthrene-d10	17.405	188	16139	0.400	ng	0.00
29) Chrysene-d12	21.587	240	11628	0.400	ng	0.00
35) Perylene-d12	24.089	264	11269	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.574	112	2496	0.152	ng	0.00
5) Phenol-d6	7.185	99	1630	0.082	ng	0.00
8) Nitrobenzene-d5	9.229	82	5302	0.431	ng	0.00
11) 2-Methylnaphthalene-d10	12.426	152	8273	0.334	ng	0.00
14) 2,4,6-Tribromophenol	16.151	330	1087	0.368	ng	0.00
15) 2-Fluorobiphenyl	13.282	172	15067	0.496	ng	-0.01
27) Fluoranthene-d10	19.434	212	15445	0.385	ng	0.00
31) Terphenyl-d14	20.015	244	13552	0.580	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.472	88	48166	6.275	ng	96
34) Bis(2-ethylhexyl)phtha...	21.443	149	2696	0.140	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091323\
Data File : BN027619.D
Acq On : 12 Sep 2023 20:43
Operator : MA/JU
Sample : 04287-10
Misc :
ALS Vial : 121 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BR-03-127-090123-3V

Quant Time: Sep 13 06:03:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 02:45:58 2023
Response via : Initial Calibration

