

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091323\
 Data File : BN027631.D
 Acq On : 13 Sep 2023 05:22
 Operator : MA/JU
 Sample : 04326-01
 Misc :
 ALS Vial : 134 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OWBR-05-135-090623

Quant Time: Sep 13 06:17:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:45:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.030	152	6160	0.400	ng	0.00
7) Naphthalene-d8	10.853	136	16167	0.400	ng	0.00
13) Acenaphthene-d10	14.662	164	7772	0.400	ng	0.00
19) Phenanthrene-d10	17.405	188	15137	0.400	ng	0.00
29) Chrysene-d12	21.587	240	10758	0.400	ng	0.00
35) Perylene-d12	24.089	264	9843	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.574	112	2490	0.155	ng	0.00
5) Phenol-d6	7.185	99	1571	0.080	ng	0.00
8) Nitrobenzene-d5	9.219	82	5425	0.460	ng	-0.01
11) 2-Methylnaphthalene-d10	12.426	152	8028	0.338	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	986	0.343	ng	0.00
15) 2-Fluorobiphenyl	13.283	172	13835	0.468	ng	-0.01
27) Fluoranthene-d10	19.430	212	15998	0.425	ng	0.00
31) Terphenyl-d14	20.015	244	14657	0.678	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.465	88	253700	33.737	ng	Qvalue 97
34) Bis(2-ethylhexyl)phtha...	21.444	149	1094	0.061	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091323\
Data File : BN027631.D
Acq On : 13 Sep 2023 05:22
Operator : MA/JU
Sample : 04326-01
Misc :
ALS Vial : 134 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OWBR-05-135-090623

Quant Time: Sep 13 06:17:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 02:45:58 2023
Response via : Initial Calibration

