

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091323\
 Data File : BN027635.D
 Acq On : 13 Sep 2023 07:50
 Operator : MA/JU
 Sample : 04327-02
 Misc :
 ALS Vial : 138 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-13A-10.5-090623

Quant Time: Sep 13 08:23:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:45:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.030	152	6765	0.400	ng	0.00
7) Naphthalene-d8	10.853	136	17961	0.400	ng	# 0.00
13) Acenaphthene-d10	14.651	164	8679	0.400	ng	-0.01
19) Phenanthrene-d10	17.406	188	17169	0.400	ng	# 0.00
29) Chrysene-d12	21.587	240	12490	0.400	ng	0.00
35) Perylene-d12	24.089	264	12192	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.567	112	2511	0.142	ng	0.00
5) Phenol-d6	7.185	99	1597	0.074	ng	0.00
8) Nitrobenzene-d5	9.219	82	5337	0.408	ng	-0.01
11) 2-Methylnaphthalene-d10	12.423	152	8175	0.310	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	1143	0.356	ng	0.00
15) 2-Fluorobiphenyl	13.283	172	14524	0.440	ng	-0.01
27) Fluoranthene-d10	19.430	212	17040	0.399	ng	0.00
31) Terphenyl-d14	20.015	244	14289	0.570	ng	0.00
Target Compounds						
17) Acenaphthene	14.715	154	14510	0.551	ng	Qvalue 98
18) Fluorene	15.693	166	3640	0.104	ng	94
28) Fluoranthene	19.463	202	1422	0.024	ng	# 92
30) Pyrene	19.825	202	1051	0.021	ng	98
34) Bis(2-ethylhexyl)phtha...	21.444	149	2542	0.123	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091323\
Data File : BN027635.D
Acq On : 13 Sep 2023 07:50
Operator : MA/JU
Sample : 04327-02
Misc :
ALS Vial : 138 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-13A-10.5-090623

Quant Time: Sep 13 08:23:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 02:45:58 2023
Response via : Initial Calibration

