

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091323\
 Data File : BN027606.D
 Acq On : 12 Sep 2023 11:56
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 108 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 12 13:03:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:45:58 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	141	0.00
2	1,4-Dioxane	0.488	0.522	-7.0	142	0.00
3	n-Nitrosodimethylamine	0.518	0.556	-7.3	154#	0.00
4 S	2-Fluorophenol	1.043	1.144	-9.7	159#	0.00
5 S	Phenol-d6	1.268	1.293	-2.0	147	0.00
6	bis(2-Chloroethyl)ether	1.226	1.255	-2.4	143	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	144	0.00
8 S	Nitrobenzene-d5	0.292	0.339	-16.1	161#	0.00
9	Naphthalene	1.091	1.172	-7.4	156#	0.00
10	Hexachlorobutadiene	0.203	0.227	-11.8	149	0.00
11 SURR	2-Methylnaphthalene-d10	0.587	0.556	5.3	132	0.00
12	2-Methylnaphthalene	0.776	0.732	5.7	130	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.148	0.138	6.8	101	0.00
15 S	2-Fluorobiphenyl	1.521	1.822	-19.8	119	0.00
16	Acenaphthylene	1.821	1.864	-2.4	104	-0.01
17	Acenaphthene	1.215	1.290	-6.2	104	0.00
18	Fluorene	1.620	1.653	-2.0	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.006	0.0	100	0.00
21	4-Bromophenyl-phenylether	0.211	0.240	-13.7	100	0.00
22	Hexachlorobenzene	0.250	0.277	-10.8	95	-0.01
23	Atrazine	0.158	0.157	0.6	89	0.00
24	Pentachlorophenol	0.096	0.086	10.4	84	0.00
25	Phenanthrene	1.176	1.249	-6.2	91	0.00
26	Anthracene	0.999	1.032	-3.3	93	0.00
27 SURR	Fluoranthene-d10	0.994	0.963	3.1	82	0.00
28	Fluoranthene	1.384	1.308	5.5	81	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	69	0.00
30	Pyrene	1.590	1.834	-15.3	79	0.00
31 S	Terphenyl-d14	0.804	0.921	-14.6	83	0.00
32	Benzo(a)anthracene	1.371	1.539	-12.3	80	0.00
33	Chrysene	1.507	1.670	-10.8	74	0.00
34	Bis(2-ethylhexyl)phthalate	0.662	0.641	3.2	66	0.00
35 I	Perylene-d12	1.000	1.000	0.0	65	0.00
36	Indeno(1,2,3-cd)pyrene	1.624	1.856	-14.3	84	0.00
37	Benzo(b)fluoranthene	1.548	1.665	-7.6	75	0.00
38	Benzo(k)fluoranthene	1.591	1.663	-4.5	74	0.00
39 C	Benzo(a)pyrene	1.447	1.509	-4.3	72	0.00
40	Dibenzo(a,h)anthracene	1.273	1.511	-18.7	88	0.00
41	Benzo(g,h,i)perylene	1.476	1.678	-13.7	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4