

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091423\
 Data File : BN027652.D
 Acq On : 13 Sep 2023 19:26
 Operator : MA/JU
 Sample : 04361-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-303-20230907

Quant Time: Sep 14 05:30:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:45:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.030	152	7456	0.400	ng	0.00
7) Naphthalene-d8	10.853	136	18519	0.400	ng	0.00
13) Acenaphthene-d10	14.662	164	8255	0.400	ng	0.00
19) Phenanthrene-d10	17.405	188	16189	0.400	ng	0.00
29) Chrysene-d12	21.587	240	11439	0.400	ng	0.00
35) Perylene-d12	24.095	264	10417	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.567	112	2523	0.130	ng	0.00
5) Phenol-d6	7.185	99	1578	0.067	ng	0.00
8) Nitrobenzene-d5	9.219	82	4955	0.367	ng	-0.01
11) 2-Methylnaphthalene-d10	12.426	152	7258	0.267	ng	0.00
14) 2,4,6-Tribromophenol	16.164	330	538	0.176	ng	0.01
15) 2-Fluorobiphenyl	13.293	172	12050	0.384	ng	0.00
27) Fluoranthene-d10	19.430	212	13316	0.331	ng	0.00
31) Terphenyl-d14	20.015	244	12060	0.525	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.444	149	795	0.042	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

