




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091423\
Data File : BN027637.D
Acq On : 13 Sep 2023 10:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 14 00:05:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 14 00:05:33 2023
Response via : Initial Calibration



TIC: BN027637.D\data.ms

(70) Pentachlorophenol (C)




17.034min (0.000) 20839.64 ng

response 443859

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	67.26
264.00	61.60	62.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091423\
Data File : BN027637.D
Acq On : 13 Sep 2023 10:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 14 00:05:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 14 00:05:33 2023
Response via : Initial Calibration



TIC: BN027637.D\data.ms

(77) Benzidine

19.669min (0.000) 0.00 ng

response 2232186

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.57
183.00	13.20	11.83
0.00	0.00	0.00