




Quant Time: Sep 15 07:46:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 15 07:46:30 2023
Response via : Initial Calibration



TIC: BN027673.D\data.ms

(70) Pentachlorophenol (C)




17.028min (0.000) 18461.90 ng

response **350765**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.18
264.00	61.60	63.21
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091423\
Data File : BN027673.D
Acq On : 14 Sep 2023 10:12
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 39 Sample Multiplier: 1

Quant Time: Sep 15 07:46:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 15 07:46:30 2023
Response via : Initial Calibration



TIC: BN027673.D\data.ms

(77) Benzidine

19.663min (0.000) 0.00 ng

response 2033388

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.84
183.00	13.20	11.50
0.00	0.00	0.00