

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091423\
 Data File : BN027672.D
 Acq On : 14 Sep 2023 08:59
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 14 11:07:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:45:58 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	147	0.00
2	1,4-Dioxane	0.400	0.373	6.8	128	0.00
3	n-Nitrosodimethylamine	0.400	0.394	1.5	147	0.00
4 S	2-Fluorophenol	0.400	0.376	6.0	141	0.00
5 S	Phenol-d6	0.400	0.347	13.3	130	0.00
6	bis(2-Chloroethyl)ether	0.400	0.367	8.3	133	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	146	-0.01
8 S	Nitrobenzene-d5	0.400	0.423	-5.7	148	-0.01
9	Naphthalene	0.400	0.389	2.8	143	-0.01
10	Hexachlorobutadiene	0.400	0.387	3.3	131	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.339	15.3	120	-0.01
12	2-Methylnaphthalene	0.400	0.337	15.8	118	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.312	22.0	84	0.00
15 S	2-Fluorobiphenyl	0.400	0.430	-7.5	105	-0.01
16	Acenaphthylene	0.400	0.368	8.0	93	-0.01
17	Acenaphthene	0.400	0.399	0.3	97	-0.01
18	Fluorene	0.400	0.371	7.3	90	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	82	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.404	-1.0	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.419	-4.7	88	-0.01
22	Hexachlorobenzene	0.400	0.406	-1.5	84	-0.01
23	Atrazine	0.400	0.357	10.8	77	0.00
24	Pentachlorophenol	0.400	0.280	30.0#	64	-0.01
25	Phenanthrene	0.400	0.401	-0.3	83	-0.01
26	Anthracene	0.400	0.381	4.8	82	0.00
27 SURR	Fluoranthene-d10	0.400	0.354	11.5	72	0.00
28	Fluoranthene	0.400	0.347	13.3	71	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	65	0.00
30	Pyrene	0.400	0.427	-6.7	69	0.00
31 S	Terphenyl-d14	0.400	0.408	-2.0	70	0.00
32	Benzo(a)anthracene	0.400	0.395	1.3	67	0.00
33	Chrysene	0.400	0.401	-0.3	63	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.362	9.5	58	0.00
35 I	Perylene-d12	0.400	0.400	0.0	60	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.401	-0.3	68	-0.01
37	Benzo(b)fluoranthene	0.400	0.392	2.0	63	0.00
38	Benzo(k)fluoranthene	0.400	0.377	5.8	61	0.00
39 C	Benzo(a)pyrene	0.400	0.369	7.8	59	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.413	-3.2	70	-0.01
41	Benzo(g,h,i)perylene	0.400	0.406	-1.5	66	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4