

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
 Data File : BN033931.D
 Acq On : 15 Sep 2024 01:25
 Operator : JU/RC
 Sample : P3952-05
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B52

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/16/2024
 Supervised By :mohammad ahmed 09/17/2024

Quant Time: Sep 15 03:08:34 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Sep 15 01:52:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.485	152	3983	0.400	ng/ul	0.00
4) Naphthalene-d8	10.239	136	11107	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.128	164	5843	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.885	188	12153m	0.400	ng/ul	0.00
17) Chrysene-d12	21.098	240	9733m	0.400	ng/ul	0.00
23) Perylene-d12	23.237	264	9749m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.101	96	22326	5.452	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.845	152	5571	0.350	ng/ul	0.00
18) Fluoranthene-d10	18.925	212	13474	0.464	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.135	88	4852	1.068	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
Data File : BN033931.D
Acq On : 15 Sep 2024 01:25
Operator : JU/RC
Sample : P3952-05
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0B52

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/16/2024
Supervised By :mohammad ahmed 09/17/2024

Quant Time: Sep 15 03:08:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Sep 15 01:52:43 2024
Response via : Initial Calibration

