

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
 Data File : BN033961.D
 Acq On : 15 Sep 2024 21:15
 Operator : JU/RC
 Sample : PB163312BL
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK312

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 09/16/2024
 Supervised By :mohammad ahmed 09/17/2024

Quant Time: Sep 16 00:15:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Sep 15 01:52:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.481	152	5281	0.400	ng/u1	0.00
4) Naphthalene-d8	10.237	136	14600	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.127	164	7552	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.879	188	15501m	0.400	ng/u1	0.00
17) Chrysene-d12	21.096	240	12359m	0.400	ng/u1	0.00
23) Perylene-d12	23.235	264	13281m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.097	96	36085	6.646	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.838	152	6311	0.301	ng/u1	0.00
18) Fluoranthene-d10	18.918	212	12793	0.347	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
Data File : BN033961.D
Acq On : 15 Sep 2024 21:15
Operator : JU/RC
Sample : PB163312BL
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK312

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/16/2024
Supervised By :mohammad ahmed 09/17/2024

Quant Time: Sep 16 00:15:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Sep 15 01:52:43 2024
Response via : Initial Calibration

