

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
 Data File : BN033982.D
 Acq On : 16 Sep 2024 10:27
 Operator : JU/RC
 Sample : PB163401BS
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS401

Quant Time: Sep 16 10:58:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Sep 15 01:52:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.477	152	6563	0.400	ng/ul	0.00
4) Naphthalene-d8	10.233	136	18829	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.123	164	9787	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.877	188	20568	0.400	ng/ul	0.00
17) Chrysene-d12	21.094	240	17763	0.400	ng/ul	0.00
23) Perylene-d12	23.237	264	19932	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.097	96	5215	0.773	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.834	152	10353	0.383	ng/ul	-0.01
18) Fluoranthene-d10	18.920	212	20811	0.393	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.126	88	15140	2.022	ng/ul#	78
5) Naphthalene	10.283	128	19088	0.379	ng/ul	100
7) 2-Methylnaphthalene	11.911	142	12047	0.378	ng/ul	100
8) 1-Methylnaphthalene	12.130	142	14454	0.440	ng/ul	99
10) Acenaphthylene	13.837	152	15961	0.359	ng/ul	100
11) Acenaphthene	14.184	153	12926	0.403	ng/ul	99
12) Fluorene	15.178	166	13898	0.388	ng/ul	99
14) Pentachlorophenol	16.539	266	2442	0.377	ng/ul	100
15) Phenanthrene	16.919	178	22253	0.367	ng/ul	99
16) Anthracene	17.012	178	19446	0.357	ng/ul	100
19) Fluoranthene	18.948	202	25588	0.365	ng/ul	99
20) Pyrene	19.315	202	26905	0.362	ng/ul	99
21) Benzo(a)anthracene	21.080	228	24488	0.350	ng/ul	99
22) Chrysene	21.132	228	27310	0.387	ng/ul	99
24) Benzo(b)fluoranthene	22.602	252	27294	0.353	ng/ul	95
25) Benzo(k)fluoranthene	22.646	252	29020	0.375	ng/ul#	94
26) Benzo(a)pyrene	23.143	252	25285	0.407	ng/ul#	93
27) Indeno(1,2,3-cd)pyrene	25.351	276	34451	0.369	ng/ul	100
28) Dibenzo(a,h)anthracene	25.368	278	26671	0.368	ng/ul	97
29) Benzo(g,h,i)perylene	25.991	276	27703	0.380	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
Data File : BN033982.D
Acq On : 16 Sep 2024 10:27
Operator : JU/RC
Sample : PB163401BS
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS401

Quant Time: Sep 16 10:58:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Sep 15 01:52:43 2024
Response via : Initial Calibration

