

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
 Data File : BN033932.D
 Acq On : 15 Sep 2024 02:01
 Operator : JU/RC
 Sample : P3952-06
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B65

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/16/2024
 Supervised By :mohammad ahmed 09/17/2024

Quant Time: Sep 15 03:08:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Sep 15 01:52:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.485	152	3955	0.400	ng/ul	0.00
4) Naphthalene-d8	10.239	136	10866	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.128	164	5665	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.885	188	11864m	0.400	ng/ul	0.00
17) Chrysene-d12	21.097	240	10238m	0.400	ng/ul	0.00
23) Perylene-d12	23.240	264	10520m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.097	96	23964	5.894	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.845	152	4060	0.260	ng/ul	0.00
18) Fluoranthene-d10	18.924	212	9704m	0.318	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.131	88	387	0.086	ng/ul#	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091424\
Data File : BN033932.D
Acq On : 15 Sep 2024 02:01
Operator : JU/RC
Sample : P3952-06
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
C0B65

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/16/2024
Supervised By :mohammad ahmed 09/17/2024


Quant Time: Sep 15 03:08:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Sep 15 01:52:43 2024
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.086 ng/ul
 RT: 3.131 min Scan# 30
 Delta R.T. -0.004 min
 Lab File: BN033932.D
 Acq: 15 Sep 2024 02:01

Instrument :
 BNA_N
 ClientSampleId :
 C0B65



Tgt Ion: 88 Resp: 38

Ion	Ratio	Lower	Upper
88	100		
43	89.9	47.0	70.4
58	37.6	53.0	79.4

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/16/2024
 Supervised By :mohammad ahmed 09/17/2024

