

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091523\
 Data File : BN027696.D
 Acq On : 15 Sep 2023 15:17
 Operator : MA/JU
 Sample : PB155422BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK422

Quant Time: Sep 16 01:31:07 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 16 01:11:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.017	152	4853	0.400	ng/ul	0.00
4) Naphthalene-d8	10.834	136	16114	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.652	164	9590	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.401	188	21548	0.400	ng/ul	0.00
17) Chrysene-d12	21.589	240	18001	0.400	ng/ul	0.00
23) Perylene-d12	24.085	264	17556	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	37808	6.185	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.418	152	7814	0.320	ng/ul	0.00
18) Fluoranthene-d10	19.422	212	19103	0.347	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091523\
Data File : BN027696.D
Acq On : 15 Sep 2023 15:17
Operator : MA/JU
Sample : PB155422BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK422

Quant Time: Sep 16 01:31:07 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 16 01:11:24 2023
Response via : Initial Calibration

