

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091523\
 Data File : BN027725.D
 Acq On : 16 Sep 2023 10:51
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4265

Quant Time: Sep 18 00:50:16 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN091523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 16 01:11:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.013	152	5209	0.400	ng/ul	0.00
4) Naphthalene-d8	10.834	136	16191	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.648	164	9157	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.400	188	22423	0.400	ng/ul	0.00
17) Chrysene-d12	21.577	240	19265	0.400	ng/ul	0.00
23) Perylene-d12	24.073	264	16800	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	2614	0.398	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.412	152	9619	0.392	ng/ul	0.00
18) Fluoranthene-d10	19.417	212	24220	0.412	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.431	88	2689	0.394	ng/ul#	91
5) Naphthalene	10.883	128	18247	0.394	ng/ul	100
7) 2-Methylnaphthalene	12.484	142	11490	0.391	ng/ul	100
8) 1-Methylnaphthalene	12.704	142	11841	0.391	ng/ul	99
10) Acenaphthylene	14.375	152	15994	0.400	ng/ul	99
11) Acenaphthene	14.712	153	13982	0.404	ng/ul	100
12) Fluorene	15.698	166	16423	0.404	ng/ul	100
14) Pentachlorophenol	17.029	266	1591	0.303	ng/ul	99
15) Phenanthrene	17.443	178	28308	0.402	ng/ul	99
16) Anthracene	17.536	178	23561	0.399	ng/ul	99
19) Fluoranthene	19.450	202	32234	0.414	ng/ul	99
20) Pyrene	19.817	202	32901	0.410	ng/ul	99
21) Benzo(a)anthracene	21.559	228	28152	0.378	ng/ul	100
22) Chrysene	21.615	228	30283	0.393	ng/ul	100
24) Benzo(b)fluoranthene	23.296	252	29770	0.398	ng/ul	98
25) Benzo(k)fluoranthene	23.348	252	28809	0.397	ng/ul	99
26) Benzo(a)pyrene	23.962	252	23203	0.385	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	26.703	276	29077	0.377	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.730	278	23317	0.375	ng/ul	98
29) Benzo(g,h,i)perylene	27.515	276	25036	0.382	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091523\
Data File : BN027725.D
Acq On : 16 Sep 2023 10:51
Operator : MA/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4265

Quant Time: Sep 18 00:50:16 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN091523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 16 01:11:24 2023
Response via : Initial Calibration

