

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091718\
 Data File : BN002724.D
 Acq On : 17 Sep 2018 14:09
 Operator : JU/SJ
 Sample : J4874-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A3ZB3

Quant Time: Sep 17 16:14:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 15 04:34:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	147058	20.00	ng/ul	0.00
18) Naphthalene-d8	10.42	136	652547	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.29	164	380199	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.04	188	833936	20.00	ng/ul	0.00
77) Chrysene-d12	21.24	240	713843	20.00	ng/ul	0.00
85) Perylene-d12	23.46	264	636999	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.22	96	21075	6.61	ng/uL	0.00
5) Phenol-d5	6.84	99	125494	10.05	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.99	67	294760	38.01	ng/ul	0.00
9) 2-Chlorophenol-d4	7.19	132	328323	31.94	ng/ul	0.00
13) 4-Methylphenol-d8	8.36	113	216763	22.41	ng/ul	0.00
19) Nitrobenzene-d5	8.80	128	194883	38.55	ng/ul	0.00
22) 2-Nitrophenol-d4	9.52	143	225234	39.37	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.05	165	369679	35.80	ng/ul	0.00
29) 4-Chloroaniline-d4	10.60	131	807	0.07	ng/ul	0.04
43) Dimethylphthalate-d6	13.70	166	1254623	42.16	ng/ul	0.00
46) Acenaphthylene-d8	13.97	160	1525254	39.25	ng/ul	0.00
51) 4-Nitrophenol-d4	14.53	143	41868	8.51	ng/ul	0.01
57) Fluorene-d10	15.29	176	1092560	41.90	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.42	200	194672	37.32	ng/ul	0.00
70) Anthracene-d10	17.13	188	1650579	41.46	ng/ul	0.00
78) Pyrene-d10	19.44	212	1691678	44.91	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.32	264	1403510	41.86	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091718\
Data File : BN002724.D
Acq On : 17 Sep 2018 14:09
Operator : JU/SJ
Sample : J4874-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A3ZB3

Quant Time: Sep 17 16:14:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 15 04:34:35 2018
Response via : Initial Calibration

