


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091719\
Data File : BN007692.D
Acq On : 17 Sep 2019 09:16
Operator : HP/JU
Sample : DFTPP03
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP03

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091619MA.M
Title : SVOA CALIBRATION
Last Update : Mon Sep 16 18:09:41 2019



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.4	713496	PASS
68	69	0.00	2	1.9	19269	PASS
69	198	0.00	100	34.8	1019129	PASS
70	69	0.00	2	0.5	5240	PASS
127	198	10	80	39.7	1163925	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	2929664	PASS
199	198	5	9	6.9	203562	PASS
275	198	10	60	28.6	836629	PASS
365	198	1	100	3.1	90797	PASS
441	198	0.01	100	12.5	364864	PASS
442	198	50	100	77.3	2265770	PASS
443	442	15	24	19.7	446314	PASS

m/z	Abundance
36.00	409.0
39.00	369.0
39.90	2690.0
41.00	469.0
42.90	580.0
44.00	14369.0
49.00	1278.0
50.80	616.0
55.10	502.0
57.00	340.0
68.90	348.0
73.10	749.0
84.00	1851.0
86.00	1109.0
87.80	331.0
90.90	308.0
95.80	426.0
106.80	335.0
114.80	414.0
118.90	307.0
132.80	412.0
138.90	317.0
152.90	402.0
161.80	336.0
166.10	461.0
167.10	1297.0
181.90	576.0
191.10	373.0
193.10	318.0
194.20	308.0
207.00	2920.0
208.00	748.0
211.00	311.0
280.90	750.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

m/z	Abundance
35.90	520.0
37.00	3182.0
38.00	7544.0
39.00	46656.0
40.00	3953.0
41.10	1220.0
42.00	342.0
42.90	743.0
44.00	12550.0
45.00	1305.0
46.90	359.0
48.10	442.0
49.00	5796.0
50.00	158528.0
51.00	618432.0
52.00	32288.0
53.00	1248.0
54.90	3479.0
56.00	19024.0
57.00	47888.0
58.00	1652.0
59.00	605.0
59.90	483.0
61.00	9599.0
62.00	12111.0
63.00	28936.0
64.00	4735.0
65.00	14111.0
66.10	1540.0
67.10	1410.0
68.10	16992.0
69.00	867008.0
70.00	3885.0
70.90	562.0
72.00	528.0
73.00	6419.0
74.00	87072.0
75.00	136512.0
76.10	45728.0
77.00	956672.0
78.00	69608.0
79.00	64744.0
80.00	47528.0
81.00	70432.0
82.00	17488.0
83.00	16374.0
84.00	3231.0
85.00	12855.0
86.00	19800.0
87.00	8370.0
88.00	2902.0
89.00	2088.0
89.70	306.0
91.00	14684.0
92.00	15431.0
93.00	99912.0
94.00	7025.0
95.00	1877.0
96.00	4763.0
98.00	80904.0
99.00	63312.0
100.00	5456.0
101.00	35224.0
101.90	1605.0
103.00	11992.0
104.00	23608.0
105.00	20648.0
106.00	6931.0
107.00	278400.0
108.00	45680.0
109.10	8663.0
110.00	530880.0
111.00	77512.0
112.00	10206.0
113.00	3227.0
113.90	984.0
115.10	1766.0
116.00	14914.0
117.00	202432.0
118.00	16496.0
119.00	2662.0
120.00	3513.0
121.00	1085.0
122.00	16624.0
123.00	28360.0
124.00	12515.0
125.00	11095.0
127.00	931776.0
128.00	71448.0
129.00	353344.0
130.00	28408.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

131.00	5622.0
131.90	2617.0
132.80	1364.0
133.20	1476.0
134.00	10935.0
135.00	31064.0
136.00	11051.0
137.00	14546.0
137.90	2900.0
139.00	2636.0
140.00	4724.0
141.00	47384.0
142.00	15647.0
143.00	10701.0
144.00	3141.0
145.00	2959.0
146.00	8436.0
147.00	24368.0
148.00	50352.0
149.00	10137.0
150.10	2633.0
151.20	5250.0
151.90	4118.0
153.00	14929.0
154.00	11723.0
155.00	26912.0
156.00	38616.0
157.10	10210.0
157.90	9562.0
159.00	7124.0
160.00	14246.0
161.00	23600.0
162.00	6799.0
163.00	2168.0
164.10	2676.0
165.00	19304.0
166.00	15656.0
167.00	96080.0
168.00	48440.0
169.00	8745.0
170.10	3306.0
170.90	3881.0
172.00	9651.0
173.00	11637.0
174.00	22472.0
175.00	39408.0
176.10	12297.0
177.00	19840.0
178.00	6543.0
179.00	74968.0
180.00	49792.0
181.00	24232.0
182.00	3591.0
182.80	1952.0
184.00	6376.0
185.00	36536.0
186.00	287808.0
187.00	84592.0
188.00	9694.0
188.90	18896.0
189.90	3127.0
191.00	8642.0
192.00	21968.0
193.00	25248.0
194.00	6669.0
196.00	63072.0
198.00	2141184.0
199.00	157056.0
200.00	11050.0
201.60	11637.0
203.00	14286.0
204.00	75744.0
205.00	131392.0
206.00	584064.0
207.00	80056.0
208.00	19888.0
209.00	5717.0
210.20	9711.0
211.10	21448.0
213.00	2024.0
214.10	487.0
214.90	6007.0
216.10	11548.0
217.00	148096.0
218.00	20752.0
219.00	2057.0
221.00	101136.0
223.00	32496.0
224.00	306176.0
225.00	85096.0
226.00	9601.0
227.00	121224.0
228.00	18704.0
229.00	26200.0
230.00	4011.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

231.10	10061.0
232.00	2427.0
233.00	3157.0
234.00	7722.0
235.00	8626.0
236.00	5385.0
237.00	10617.0
238.00	1423.0
239.00	5357.0
240.00	3839.0
241.00	5837.0
242.00	14778.0
243.00	17336.0
244.00	238464.0
245.00	34608.0
246.00	43992.0
246.90	9850.0
248.10	2371.0
249.00	9593.0
250.00	1514.0
251.10	1891.0
252.00	2677.0
253.00	5618.0
255.00	1192448.0
256.00	184256.0
257.00	14368.0
258.00	66560.0
259.00	10600.0
260.10	1780.0
260.90	2048.0
261.90	318.0
262.90	922.0
264.00	1538.0
265.00	27296.0
266.00	3845.0
267.00	691.0
268.00	397.0
269.00	354.0
269.90	1008.0
270.90	3034.0
272.10	2709.0
273.00	34672.0
274.00	96416.0
275.00	547584.0
276.00	80144.0
277.00	43840.0
278.00	7557.0
279.00	1776.0
281.20	1155.0
282.00	1066.0
283.00	4980.0
283.90	3257.0
285.00	7170.0
286.00	1660.0
288.00	511.0
289.00	2119.0
290.00	1805.0
291.20	914.0
292.00	1964.0
293.00	10511.0
294.00	2479.0
296.00	139264.0
297.00	19064.0
297.90	1196.0
298.90	433.0
301.00	1704.0
301.90	2421.0
303.00	15436.0
304.00	5003.0
305.00	560.0
306.80	461.0
307.90	2359.0
308.90	1306.0
310.10	2021.0
311.20	538.0
312.10	311.0
313.00	1308.0
314.00	6597.0
315.00	14699.0
316.00	9273.0
317.00	1662.0
317.80	354.0
320.00	624.0
321.10	4320.0
322.00	2757.0
323.00	39896.0
324.00	8621.0
324.90	942.0
325.80	774.0
326.10	739.0
326.90	6911.0
328.00	4324.0
328.90	619.0
331.90	3112.0
333.00	3710.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

334.00	26040.0
335.00	7515.0
336.10	924.0
338.90	782.0
339.90	897.0
341.10	4099.0
341.90	1665.0
345.90	8690.0
346.90	1685.0
349.90	610.0
351.00	911.0
352.00	12680.0
353.00	8974.0
354.00	13347.0
355.00	3503.0
358.00	411.0
359.00	982.0
359.80	302.0
361.00	453.0
365.00	51656.0
366.00	8038.0
366.90	875.0
370.00	634.0
370.90	3094.0
372.00	21768.0
373.00	5375.0
374.00	688.0
377.00	857.0
377.90	337.0
383.00	6142.0
384.00	1719.0
385.00	620.0
390.00	3091.0
391.10	1535.0
392.00	1575.0
396.90	316.0
400.90	1400.0
402.00	9107.0
403.00	12368.0
404.00	5587.0
405.00	726.0
415.00	633.0
419.90	437.0
421.00	10283.0
422.00	10690.0
423.00	74288.0
424.00	15998.0
425.00	2195.0
439.10	425.0
441.00	185280.0
442.00	1124864.0
443.00	231360.0
444.00	21832.0
445.00	1329.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

m/z	Abundance
36.00	733.0
37.00	3267.0
38.00	9722.0
39.00	61224.0
40.00	4997.0
41.00	1880.0
43.00	607.0
44.00	10200.0
45.00	1798.0
46.90	539.0
48.00	482.0
49.10	7614.0
50.00	208896.0
51.00	811392.0
52.00	44048.0
53.00	2116.0
55.00	4596.0
56.00	25112.0
57.00	60904.0
58.00	2831.0
59.00	1196.0
59.90	1095.0
61.00	12817.0
62.00	14193.0
63.00	40696.0
64.00	6450.0
65.00	21584.0
65.90	1921.0
67.10	1706.0
68.10	21280.0
69.00	1165312.0
70.00	5761.0
71.00	1075.0
71.90	696.0
73.00	9088.0
74.00	109768.0
75.00	179392.0
76.10	62768.0
77.00	1314304.0
78.00	91064.0
79.00	88408.0
80.00	66192.0
81.00	95456.0
82.00	23080.0
83.00	22384.0
84.00	3397.0
85.00	17736.0
86.00	26832.0
87.00	11532.0
88.00	4279.0
89.00	2550.0
90.00	847.0
91.00	19320.0
92.00	24424.0
93.00	141504.0
94.00	9790.0
95.00	2889.0
96.00	7394.0
97.20	3454.0
98.00	109200.0
99.00	84000.0
100.00	7888.0
101.00	49824.0
102.10	2593.0
103.00	16536.0
104.00	32624.0
105.00	31288.0
106.00	10401.0
107.00	391232.0
108.00	63256.0
109.10	12511.0
110.00	738944.0
111.00	108656.0
112.00	11719.0
113.00	4234.0
114.10	844.0
115.10	2183.0
116.00	21456.0
117.00	291200.0
118.00	22632.0
119.00	3307.0
119.90	5264.0
121.10	2129.0
122.00	24352.0
123.00	37304.0
124.00	16440.0
125.00	16326.0
127.00	1334272.0
128.00	106296.0
129.00	491712.0
130.00	42776.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

131.00	8550.0
132.00	3877.0
132.90	2423.0
134.00	14783.0
135.00	41832.0
136.00	17440.0
137.00	22264.0
138.10	4223.0
139.00	3026.0
140.00	6930.0
141.00	69088.0
142.00	21432.0
143.00	14609.0
144.00	3911.0
145.00	4378.0
146.00	12627.0
147.00	33008.0
148.00	76696.0
149.00	16028.0
150.00	4150.0
151.20	8478.0
151.80	5520.0
153.00	22864.0
154.00	16456.0
155.00	41280.0
156.00	58160.0
157.00	12001.0
157.90	14673.0
159.00	9852.0
160.00	21408.0
161.00	32832.0
162.00	9318.0
163.10	2449.0
164.10	4093.0
165.00	29336.0
166.00	21264.0
167.00	144640.0
168.00	63816.0
169.00	12775.0
170.00	4378.0
171.10	6748.0
172.00	13470.0
173.00	18360.0
174.00	30368.0
175.00	58008.0
176.10	17544.0
177.00	29024.0
178.00	9817.0
179.00	115560.0
180.00	77632.0
181.00	35232.0
182.00	5980.0
183.00	3732.0
184.00	9307.0
185.00	55656.0
186.00	452608.0
187.00	127184.0
188.00	13498.0
189.00	28728.0
190.00	4700.0
191.00	13181.0
192.00	35720.0
193.00	38880.0
194.00	8810.0
195.10	5807.0
196.00	95280.0
198.00	3340288.0
199.00	230336.0
200.00	16616.0
201.40	14471.0
203.00	23656.0
204.00	116704.0
205.00	206016.0
206.00	911104.0
207.00	121608.0
208.00	27520.0
209.00	9134.0
211.00	34504.0
213.00	2764.0
214.00	1192.0
215.00	10487.0
216.00	19096.0
217.00	231360.0
218.00	31040.0
219.00	2765.0
221.00	165056.0
223.00	53024.0
224.00	516288.0
225.00	136960.0
226.00	13182.0
227.00	202432.0
228.00	28344.0
229.00	42952.0
230.00	6071.0
231.00	19792.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

232.00	2876.0
233.00	3349.0
234.00	11356.0
235.00	14950.0
236.00	8872.0
237.00	16022.0
237.90	2575.0
239.00	7999.0
240.10	5699.0
241.00	10644.0
242.00	26032.0
243.10	29240.0
244.00	413056.0
245.00	53152.0
246.00	69536.0
247.00	15823.0
248.00	3098.0
249.00	13519.0
250.00	2330.0
250.90	3260.0
252.00	3749.0
253.00	8073.0
255.00	2063360.0
256.00	307712.0
257.00	24200.0
258.00	110760.0
259.00	17800.0
260.00	3257.0
261.00	3341.0
262.00	772.0
263.00	1386.0
264.10	2684.0
265.00	41928.0
266.00	5609.0
266.90	909.0
269.10	518.0
269.90	1728.0
271.00	3420.0
272.00	5540.0
273.00	58928.0
274.00	157120.0
275.00	974656.0
276.00	130800.0
277.00	74224.0
278.00	11972.0
279.00	2491.0
280.10	457.0
280.90	784.0
282.00	1819.0
283.00	7641.0
284.10	5103.0
285.00	13391.0
286.00	2760.0
287.10	573.0
287.90	1115.0
289.00	2917.0
290.00	2278.0
291.00	1402.0
292.00	3763.0
293.00	16920.0
294.00	4924.0
296.00	234624.0
297.00	35096.0
298.00	2243.0
298.90	768.0
300.00	373.0
300.90	3449.0
302.00	3910.0
303.00	27280.0
304.00	8049.0
305.00	1109.0
305.90	366.0
306.90	465.0
308.00	3369.0
308.90	2497.0
310.00	3412.0
311.00	1066.0
312.10	749.0
313.00	2136.0
314.00	11361.0
315.00	25432.0
316.00	14983.0
317.00	2798.0
318.00	481.0
319.00	674.0
319.90	816.0
321.00	7626.0
321.90	4296.0
323.00	79528.0
324.00	16095.0
325.00	1551.0
326.10	1554.0
327.00	15450.0
328.00	7576.0
328.90	1454.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

330.00	331.0
330.90	462.0
332.00	6158.0
333.00	6338.0
334.00	49224.0
335.00	12497.0
336.10	2019.0
339.10	1432.0
340.00	1226.0
341.00	9232.0
341.90	2858.0
343.10	898.0
346.00	16049.0
347.00	3447.0
347.90	502.0
350.10	1018.0
351.10	1896.0
352.00	23368.0
353.00	16267.0
354.00	25088.0
355.00	5211.0
356.00	556.0
358.00	581.0
359.00	1683.0
359.90	482.0
361.10	315.0
365.00	107232.0
366.00	14476.0
366.90	1294.0
370.00	2142.0
371.00	5373.0
372.00	42080.0
373.00	10908.0
373.90	1373.0
376.90	1259.0
382.10	443.0
382.90	9986.0
384.00	3572.0
384.90	927.0
389.10	468.0
390.00	5402.0
390.90	3736.0
392.00	2600.0
393.00	491.0
394.90	580.0
396.00	432.0
397.10	404.0
401.00	2819.0
402.00	17480.0
403.00	23440.0
404.00	7847.0
405.00	1282.0
410.00	805.0
415.00	1300.0
421.00	20176.0
422.00	21408.0
423.00	143808.0
424.00	30648.0
425.00	3303.0
426.00	354.0
441.00	406336.0
442.00	2522624.0
443.00	505152.0
444.00	45560.0
445.00	2362.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

m/z	Abundance
36.00	506.0
37.10	2572.0
38.00	8566.0
39.10	51688.0
40.00	4136.0
41.00	1488.0
43.10	656.0
44.00	10870.0
45.00	1650.0
47.00	368.0
48.20	588.0
49.10	6130.0
50.00	179008.0
51.00	712512.0
52.00	36624.0
52.90	1701.0
53.90	312.0
55.00	4423.0
56.00	24104.0
57.00	56488.0
58.00	2071.0
59.00	588.0
59.90	502.0
61.00	10003.0
62.00	10992.0
63.00	36408.0
64.00	5019.0
65.00	18128.0
66.00	1423.0
67.10	1547.0
68.10	19536.0
69.00	1026112.0
70.00	6076.0
71.10	680.0
72.10	407.0
73.00	8285.0
74.00	100176.0
75.00	163648.0
76.10	54072.0
77.00	1178112.0
78.00	79248.0
79.00	79168.0
80.00	55456.0
81.00	78600.0
82.00	21336.0
83.00	19560.0
84.00	3276.0
85.00	13963.0
86.00	23600.0
87.00	10809.0
88.00	3661.0
89.00	2612.0
89.80	528.0
91.00	19384.0
92.00	21392.0
93.00	125656.0
94.00	8351.0
95.00	2432.0
96.00	6367.0
98.00	100472.0
99.00	78488.0
100.00	6304.0
101.00	48488.0
101.90	2144.0
103.00	16102.0
104.00	31720.0
105.00	26592.0
106.00	9197.0
107.00	353792.0
108.00	57720.0
109.10	10491.0
110.00	673216.0
111.00	98336.0
112.00	11989.0
112.90	3909.0
116.00	20088.0
117.00	262848.0
118.00	20120.0
119.00	2405.0
120.10	3888.0
121.00	1473.0
122.00	22776.0
123.00	35080.0
124.00	15354.0
125.00	14012.0
127.00	1225728.0
128.00	94048.0
129.00	467392.0
130.00	36880.0
131.00	7901.0
132.00	4050.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

132.80	1713.0
134.00	12381.0
135.00	37472.0
136.00	15016.0
137.00	20296.0
137.90	3961.0
139.00	2813.0
140.00	5383.0
141.00	61608.0
142.00	22496.0
143.00	14144.0
144.10	4199.0
145.10	3275.0
145.90	11261.0
147.00	30672.0
148.00	69768.0
149.00	14465.0
150.00	4168.0
151.10	8711.0
151.90	5345.0
153.00	20904.0
154.00	16544.0
155.00	38288.0
156.00	57000.0
157.00	11792.0
158.00	11199.0
159.00	9414.0
160.00	20528.0
161.00	31656.0
161.90	8186.0
163.00	2678.0
164.00	3859.0
165.00	24192.0
166.00	22000.0
167.00	128720.0
168.00	59696.0
169.00	10605.0
170.00	4620.0
170.90	5923.0
172.00	12851.0
173.00	16218.0
174.00	30200.0
175.00	54240.0
175.90	15306.0
177.00	25576.0
178.00	8691.0
179.00	107296.0
180.00	71896.0
181.00	34416.0
182.00	4635.0
182.90	3785.0
184.00	8478.0
185.00	54760.0
186.00	424384.0
187.00	122704.0
188.10	12810.0
189.00	27352.0
190.00	3987.0
191.00	11800.0
192.00	34608.0
193.00	37256.0
194.00	7893.0
195.00	5440.0
196.00	94576.0
198.00	3307520.0
199.00	223296.0
200.00	15390.0
201.50	16199.0
203.00	22224.0
204.00	115512.0
205.00	199424.0
206.00	857344.0
207.00	117816.0
208.00	28760.0
209.00	8573.0
211.00	34456.0
212.90	2522.0
213.90	1125.0
215.00	10195.0
216.00	19936.0
217.00	227008.0
218.00	29832.0
219.00	2924.0
221.00	158528.0
221.80	33328.0
223.00	54360.0
224.00	503360.0
225.00	124200.0
226.10	13759.0
227.00	193856.0
228.00	28192.0
229.00	42504.0
230.00	5803.0
231.00	17920.0
231.90	2890.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

233.00	3228.0
234.00	13206.0
235.00	16800.0
236.00	9483.0
237.00	16672.0
238.00	2045.0
239.00	7107.0
240.00	5059.0
240.90	11674.0
242.00	26528.0
243.00	29232.0
244.00	389440.0
245.00	52024.0
246.00	72352.0
247.00	15742.0
248.00	3854.0
249.00	15768.0
250.00	2699.0
250.90	3411.0
252.00	3645.0
253.00	7749.0
255.00	2060288.0
256.00	297600.0
257.00	21608.0
258.00	108240.0
259.00	17648.0
260.00	3693.0
261.00	3756.0
261.90	476.0
263.00	1385.0
264.10	2906.0
265.00	43504.0
266.00	6366.0
267.00	795.0
268.00	631.0
269.10	524.0
270.00	2399.0
271.00	3678.0
272.10	4867.0
273.00	62672.0
274.00	167040.0
275.00	987648.0
276.00	125616.0
277.00	73432.0
278.00	13449.0
279.00	2431.0
280.10	417.0
280.90	992.0
282.00	2218.0
283.00	7554.0
284.10	5305.0
285.00	12773.0
286.00	2876.0
286.90	322.0
288.00	824.0
289.00	3526.0
289.90	2190.0
291.00	1990.0
292.00	3615.0
293.00	17920.0
294.00	4596.0
295.00	5591.0
296.00	235264.0
297.00	30840.0
298.00	2129.0
299.10	758.0
300.90	3491.0
302.00	4052.0
303.00	29392.0
304.00	7525.0
305.00	1113.0
306.00	508.0
307.00	515.0
308.00	3993.0
309.00	2283.0
310.00	3134.0
311.10	1349.0
311.80	762.0
313.00	2065.0
314.00	13011.0
315.00	25448.0
316.00	16456.0
317.00	2791.0
318.00	321.0
318.90	568.0
320.10	1247.0
321.00	9791.0
322.00	4457.0
323.00	81216.0
324.00	14877.0
325.00	1038.0
326.10	1525.0
327.00	16102.0
328.00	7217.0
329.00	1689.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

329.90	424.0
331.00	606.0
332.00	6113.0
333.00	7123.0
334.00	56672.0
335.00	13529.0
336.00	1608.0
338.90	1126.0
340.00	1421.0
341.00	9643.0
342.00	3038.0
342.90	383.0
346.00	17952.0
347.00	3527.0
347.90	556.0
350.00	831.0
350.90	1578.0
352.00	24432.0
353.00	20520.0
354.00	27456.0
355.00	5418.0
356.10	829.0
356.90	422.0
357.90	520.0
359.00	2173.0
360.10	498.0
361.10	466.0
363.10	491.0
365.00	113504.0
366.00	16560.0
367.00	1111.0
370.00	2481.0
371.00	7100.0
372.00	47408.0
373.00	12280.0
374.00	1244.0
377.00	1201.0
378.10	476.0
382.00	495.0
383.00	13396.0
384.00	3231.0
384.90	1217.0
385.90	338.0
390.00	6889.0
391.00	4658.0
392.10	3540.0
392.90	672.0
394.90	473.0
396.10	514.0
397.00	643.0
400.90	2186.0
402.00	19616.0
403.00	28536.0
404.00	9969.0
405.00	2449.0
410.00	995.0
415.00	1074.0
416.00	575.0
421.00	24752.0
422.00	24160.0
423.00	181184.0
424.00	37008.0
424.90	4426.0
426.00	443.0
439.20	635.0
441.00	502976.0
442.00	3149824.0
443.00	602432.0
444.00	53968.0
445.00	3045.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03