

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
 Data File : BN007711.D
 Acq On : 18 Sep 2019 18:29
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN091819

Quant Time: Sep 18 19:29:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.446	0.362	18.8	100	0.00
3	n-Nitrosodimethylamine	0.204	0.316	-54.9#	106	0.00
4 S	2-Fluorophenol	1.377	1.162	15.6	101	0.00
5 S	Phenol-d6	1.962	1.494	23.9	100	0.00
6	bis(2-Chloroethyl)ether	1.133	1.020	10.0	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.350	0.329	6.0	104	0.00
9	Naphthalene	1.123	1.114	0.8	107	0.00
10	Hexachlorobutadiene	0.236	0.234	0.8	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.674	0.653	3.1	107	0.00
12	2-Methylnaphthalene	0.759	0.749	1.3	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.218	0.193	11.5	115	0.00
15 S	2-Fluorobiphenyl	1.647	1.538	6.6	110	0.00
16	Acenaphthylene	2.121	1.889	10.9	108	-0.01
17	Acenaphthene	1.368	1.231	10.0	107	0.00
18	Fluorene	1.753	1.595	9.0	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.243	0.243	0.0	109	0.00
21	Hexachlorobenzene	0.265	0.270	-1.9	110	0.00
22	Pentachlorophenol	0.098	0.080	18.4	100	0.00
23	Phenanthrene	1.246	1.242	0.3	108	0.00
24	Anthracene	1.125	1.079	4.1	106	0.00
25 SURR	Fluoranthene-d10	1.348	1.314	2.5	105	0.00
26	Fluoranthene	1.549	1.523	1.7	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pvrene	1.364	1.348	1.2	105	0.00
29 S	Terphenyl-d14	0.983	0.964	1.9	104	0.00
30	Benzo(a)anthracene	1.473	1.426	3.2	103	0.00
31	Chrysene	1.436	1.419	1.2	104	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.600	20.8	90	0.00
33	Indeno(1,2,3-cd)pyrene	1.797	1.651	8.1	96	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.389	1.354	2.5	101	0.00
36	Benzo(k)fluoranthene	1.373	1.321	3.8	98	0.00
37 C	Benzo(a)pyrene	1.305	1.232	5.6	97	0.00
38	Dibenzo(a,h)anthracene	1.335	1.261	5.5	96	0.00
39	Benzo(a,h,i)perylene	1.321	1.245	5.8	96	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
Data File : BN007711.D
Acq On : 18 Sep 2019 18:29
Operator : HP/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN091819

Quant Time: Sep 18 19:29:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			