

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
 Data File : BN007713.D
 Acq On : 18 Sep 2019 19:41
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 07:15:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.446	0.363	18.6	100	0.00
3	n-Nitrosodimethylamine	0.204	0.323	-58.3#	108	0.00
4 S	2-Fluorophenol	1.377	1.162	15.6	101	0.00
5 S	Phenol-d6	1.962	1.526	22.2	102	0.00
6	bis(2-Chloroethyl)ether	1.133	0.959	15.4	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.350	0.396	-13.1	124	0.00
9	Naphthalene	1.123	1.126	-0.3	107	0.00
10	Hexachlorobutadiene	0.236	0.234	0.8	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.674	0.644	4.5	105	0.00
12	2-Methylnaphthalene	0.759	0.751	1.1	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.218	0.167	23.4	99	0.00
15 S	2-Fluorobiphenyl	1.647	1.893	-14.9	133	0.00
16	Acenaphthylene	2.121	1.863	12.2	105	0.00
17	Acenaphthene	1.368	1.227	10.3	106	0.00
18	Fluorene	1.753	1.589	9.4	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.243	0.238	2.1	104	0.00
21	Hexachlorobenzene	0.265	0.270	-1.9	108	0.00
22	Pentachlorophenol	0.098	0.078	20.4	96	0.00
23	Phenanthrene	1.246	1.232	1.1	105	0.00
24	Anthracene	1.125	1.061	5.7	102	0.00
25 SURR	Fluoranthene-d10	1.348	1.255	6.9	99	0.00
26	Fluoranthene	1.549	1.496	3.4	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pvrene	1.364	1.361	0.2	102	0.00
29 S	Terphenyl-d14	0.983	1.243	-26.4#	129	0.00
30	Benzo(a)anthracene	1.473	1.446	1.8	100	0.00
31	Chrysene	1.436	1.403	2.3	98	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.580	23.5	84	-0.01
33	Indeno(1,2,3-cd)pyrene	1.797	1.718	4.4	96	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.389	1.325	4.6	97	0.00
36	Benzo(k)fluoranthene	1.373	1.377	-0.3	100	0.00
37 C	Benzo(a)pyrene	1.305	1.237	5.2	95	0.00
38	Dibenzo(a,h)anthracene	1.335	1.279	4.2	96	0.00
39	Benzo(a,h,i)perylene	1.321	1.246	5.7	94	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
Data File : BN007713.D
Acq On : 18 Sep 2019 19:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 19 07:15:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			