

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
 Data File : BN007730.D
 Acq On : 19 Sep 2019 05:50
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 07:15:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.371	7.3	123	0.00
3	n-Nitrosodimethylamine	0.400	0.573	-43.2#	106	0.00
4 S	2-Fluorophenol	0.400	0.318	20.5	103	0.00
5 S	Phenol-d6	0.400	0.327	18.3	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	122	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.383	4.3	110	0.00
9	Naphthalene	0.400	0.402	-0.5	113	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	111	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	113	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.287	28.3#	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	118	0.00
16	Acenaphthylene	0.400	0.347	13.3	103	-0.01
17	Acenaphthene	0.400	0.359	10.3	105	0.00
18	Fluorene	0.400	0.359	10.3	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	103	-0.01
21	Hexachlorobenzene	0.400	0.416	-4.0	108	-0.01
22	Pentachlorophenol	0.400	0.310	22.5	91	-0.01
23	Phenanthrene	0.400	0.401	-0.3	104	0.00
24	Anthracene	0.400	0.372	7.0	99	0.00
25 SURR	Fluoranthene-d10	0.400	0.376	6.0	98	0.00
26	Fluoranthene	0.400	0.387	3.3	101	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	102	-0.01
28	Pvrene	0.400	0.382	4.5	100	0.00
29 S	Terphenyl-d14	0.400	0.419	-4.7	110	0.00
30	Benzo(a)anthracene	0.400	0.360	10.0	95	0.00
31	Chrysene	0.400	0.407	-1.7	106	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.234	41.5#	66	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.398	0.5	103	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	100	-0.01
35	Benzo(b)fluoranthene	0.400	0.373	6.8	99	0.00
36	Benzo(k)fluoranthene	0.400	0.411	-2.7	107	0.00
37 C	Benzo(a)pyrene	0.400	0.364	9.0	96	0.00
38	Dibenzo(a,h)anthracene	0.400	0.392	2.0	102	0.00
39	Benzo(a,h,i)perylene	0.400	0.379	5.3	99	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
Data File : BN007730.D
Acq On : 19 Sep 2019 05:50
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 19 07:15:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				