

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
 Data File : BN007730.D
 Acq On : 19 Sep 2019 05:50
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 07:15:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.446	0.414	7.2	123	0.00
3	n-Nitrosodimethylamine	0.204	0.292	-43.1#	106	0.00
4 S	2-Fluorophenol	1.377	1.095	20.5	103	0.00
5 S	Phenol-d6	1.962	1.406	28.3#	102	0.00
6	bis(2-Chloroethyl)ether	1.133	1.080	4.7	122	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.350	0.335	4.3	110	0.00
9	Naphthalene	1.123	1.130	-0.6	113	0.00
10	Hexachlorobutadiene	0.236	0.235	0.4	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.674	0.649	3.7	111	0.00
12	2-Methylnaphthalene	0.759	0.748	1.4	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.218	0.156	28.4#	92	0.00
15 S	2-Fluorobiphenyl	1.647	1.687	-2.4	118	0.00
16	Acenaphthylene	2.121	1.842	13.2	103	-0.01
17	Acenaphthene	1.368	1.227	10.3	105	0.00
18	Fluorene	1.753	1.572	10.3	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.243	0.239	1.6	103	-0.01
21	Hexachlorobenzene	0.265	0.276	-4.2	108	-0.01
22	Pentachlorophenol	0.098	0.076	22.4	91	-0.01
23	Phenanthrene	1.246	1.251	-0.4	104	0.00
24	Anthracene	1.125	1.046	7.0	99	0.00
25 SURR	Fluoranthene-d10	1.348	1.266	6.1	98	0.00
26	Fluoranthene	1.549	1.499	3.2	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
28	Pvrene	1.364	1.301	4.6	100	0.00
29 S	Terphenyl-d14	0.983	1.029	-4.7	110	0.00
30	Benzo(a)anthracene	1.473	1.327	9.9	95	0.00
31	Chrysene	1.436	1.462	-1.8	106	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.443	41.6#	66	-0.01
33	Indeno(1,2,3-cd)pyrene	1.797	1.787	0.6	103	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.01
35	Benzo(b)fluoranthene	1.389	1.294	6.8	99	0.00
36	Benzo(k)fluoranthene	1.373	1.411	-2.8	107	0.00
37 C	Benzo(a)pyrene	1.305	1.189	8.9	96	0.00
38	Dibenzo(a,h)anthracene	1.335	1.308	2.0	102	0.00
39	Benzo(a,h,i)perylene	1.321	1.253	5.1	99	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
Data File : BN007730.D
Acq On : 19 Sep 2019 05:50
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 19 07:15:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			