

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
 Data File : BN007732.D
 Acq On : 19 Sep 2019 07:39
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 15:10:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.446	0.414	7.2	132	0.00
3	n-Nitrosodimethylamine	0.204	0.276	-35.3#	107	0.00
4 S	2-Fluorophenol	1.377	1.064	22.7	107	0.00
5 S	Phenol-d6	1.962	1.405	28.4#	109	0.00
6	bis(2-Chloroethyl)ether	1.133	1.056	6.8	128	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.350	0.327	6.6	118	0.00
9	Naphthalene	1.123	1.128	-0.4	124	0.00
10	Hexachlorobutadiene	0.236	0.232	1.7	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.674	0.662	1.8	124	0.00
12	2-Methylnaphthalene	0.759	0.755	0.5	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.218	0.156	28.4#	103	0.00
15 S	2-Fluorobiphenyl	1.647	1.690	-2.6	133	0.00
16	Acenaphthylene	2.121	1.840	13.2	116	-0.01
17	Acenaphthene	1.368	1.228	10.2	118	-0.01
18	Fluorene	1.753	1.595	9.0	120	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.243	0.240	1.2	119	-0.01
21	Hexachlorobenzene	0.265	0.274	-3.4	123	0.00
22	Pentachlorophenol	0.098	0.071	27.6#	99	0.00
23	Phenanthrene	1.246	1.248	-0.2	120	0.00
24	Anthracene	1.125	1.042	7.4	114	0.00
25 SURR	Fluoranthene-d10	1.348	1.246	7.6	111	0.00
26	Fluoranthene	1.549	1.474	4.8	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
28	Pvrene	1.364	1.346	1.3	112	0.00
29 S	Terphenyl-d14	0.983	1.067	-8.5	123	0.00
30	Benzo(a)anthracene	1.473	1.359	7.7	105	0.00
31	Chrysene	1.436	1.430	0.4	112	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.443	41.6#	71	-0.01
33	Indeno(1,2,3-cd)pyrene	1.797	1.707	5.0	106	0.00
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.389	1.308	5.8	104	0.00
36	Benzo(k)fluoranthene	1.373	1.432	-4.3	114	0.00
37 C	Benzo(a)pyrene	1.305	1.194	8.5	101	0.00
38	Dibenzo(a,h)anthracene	1.335	1.284	3.8	105	0.00
39	Benzo(a,h,i)perylene	1.321	1.228	7.0	101	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
Data File : BN007732.D
Acq On : 19 Sep 2019 07:39
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 19 15:10:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			