

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
 Data File : BN007739.D
 Acq On : 19 Sep 2019 14:01
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 15:22:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2	1,4-Dioxane	0.446	0.410	8.1	146	0.00
3	n-Nitrosodimethylamine	0.204	0.295	-44.6#	128	0.00
4 S	2-Fluorophenol	1.377	1.055	23.4	119	0.00
5 S	Phenol-d6	1.962	1.372	30.1#	119	0.00
6	bis(2-Chloroethyl)ether	1.133	1.064	6.1	144	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
8 S	Nitrobenzene-d5	0.350	0.379	-8.3	150	0.00
9	Naphthalene	1.123	1.139	-1.4	137	0.00
10	Hexachlorobutadiene	0.236	0.233	1.3	132	0.00
11 SURR	2-Methylnaphthalene-d10	0.674	0.644	4.5	132	0.00
12	2-Methylnaphthalene	0.759	0.754	0.7	137	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.218	0.148	32.1#	105	0.00
15 S	2-Fluorobiphenyl	1.647	1.953	-18.6	165#	0.00
16	Acenaphthylene	2.121	1.810	14.7	123	-0.01
17	Acenaphthene	1.368	1.218	11.0	126	-0.01
18	Fluorene	1.753	1.582	9.8	127	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
20	4-Bromophenyl-phenylether	0.243	0.238	2.1	125	0.00
21	Hexachlorobenzene	0.265	0.272	-2.6	130	-0.01
22	Pentachlorophenol	0.098	0.066	32.7#	96	-0.01
23	Phenanthrene	1.246	1.242	0.3	127	0.00
24	Anthracene	1.125	1.035	8.0	120	0.00
25 SURR	Fluoranthene-d10	1.348	1.219	9.6	115	0.00
26	Fluoranthene	1.549	1.475	4.8	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
28	Pvrene	1.364	1.330	2.5	120	0.00
29 S	Terphenyl-d14	0.983	1.228	-24.9	153#	0.00
30	Benzo(a)anthracene	1.473	1.336	9.3	112	0.00
31	Chrysene	1.436	1.447	-0.8	122	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.404	46.7#	71	0.00
33	Indeno(1,2,3-cd)pyrene	1.797	1.782	0.8	120	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.389	1.322	4.8	116	0.00
36	Benzo(k)fluoranthene	1.373	1.429	-4.1	124	0.00
37 C	Benzo(a)pyrene	1.305	1.182	9.4	109	0.00
38	Dibenzo(a,h)anthracene	1.335	1.320	1.1	119	0.00
39	Benzo(a,h,i)perylene	1.321	1.239	6.2	112	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091819\
Data File : BN007739.D
Acq On : 19 Sep 2019 14:01
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 19 15:22:31 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			