

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091823\
 Data File : BN027741.D
 Acq On : 18 Sep 2023 19:19
 Operator : MA/JU
 Sample : 04437-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW10A-GW-280-282

Quant Time: Sep 19 04:03:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 18 14:40:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.015	152	2228	0.400	ng	0.00
7) Naphthalene-d8	10.831	136	7735	0.400	ng	0.00
13) Acenaphthene-d10	14.640	164	5601	0.400	ng	0.00
19) Phenanthrene-d10	17.393	188	14859	0.400	ng	# 0.00
29) Chrysene-d12	21.578	240	17927	0.400	ng	0.00
35) Perylene-d12	24.077	264	22526	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.545	112	355	0.061	ng	0.00
5) Phenol-d6	7.163	99	307	0.043	ng	0.00
8) Nitrobenzene-d5	9.208	82	1054	0.168	ng	0.00
11) 2-Methylnaphthalene-d10	12.411	152	1793	0.157	ng	0.00
14) 2,4,6-Tribromophenol	16.139	330	309	0.118	ng	0.00
15) 2-Fluorobiphenyl	13.272	172	3605	0.189	ng	0.00
27) Fluoranthene-d10	19.420	212	7205	0.173	ng	0.00
31) Terphenyl-d14	20.006	244	6735	0.200	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.429	88	506	0.194	ng	# 74
9) Naphthalene	10.884	128	1254	0.059	ng	# 72
25) Phenanthrene	17.442	178	944	0.022	ng	# 82
34) Bis(2-ethylhexyl)phtha...	21.434	149	66615	2.207	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091823\
Data File : BN027741.D
Acq On : 18 Sep 2023 19:19
Operator : MA/JU
Sample : 04437-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW10A-GW-280-282

Quant Time: Sep 19 04:03:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 18 14:40:21 2023
Response via : Initial Calibration

