

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091823\
 Data File : BN027742.D
 Acq On : 18 Sep 2023 19:56
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 04:03:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 18 14:40:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	113	0.00
2	1,4-Dioxane	0.400	0.382	4.5	97	0.00
3	n-Nitrosodimethylamine	0.400	0.357	10.8	93	0.00
4 S	2-Fluorophenol	0.400	0.385	3.8	105	0.00
5 S	Phenol-d6	0.400	0.384	4.0	109	0.00
6	bis(2-Chloroethyl)ether	0.400	0.388	3.0	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	117	0.00
8 S	Nitrobenzene-d5	0.400	0.346	13.5	99	0.00
9	Naphthalene	0.400	0.395	1.3	108	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	107	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.01
14 S	2,4,6-Tribromophenol	0.400	0.378	5.5	117	0.00
15 S	2-Fluorobiphenyl	0.400	0.365	8.8	110	0.00
16	Acenaphthylene	0.400	0.377	5.8	112	0.00
17	Acenaphthene	0.400	0.395	1.3	111	0.00
18	Fluorene	0.400	0.396	1.0	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.319	20.3	90	0.00
21	4-Bromophenyl-phenylether	0.400	0.371	7.3	105	0.01
22	Hexachlorobenzene	0.400	0.385	3.8	107	0.00
23	Atrazine	0.400	0.386	3.5	111	0.00
24	Pentachlorophenol	0.400	0.326	18.5	103	0.01
25	Phenanthrene	0.400	0.394	1.5	110	0.01
26	Anthracene	0.400	0.386	3.5	110	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	108	0.00
28	Fluoranthene	0.400	0.386	3.5	108	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	115	0.00
30	Pyrene	0.400	0.403	-0.8	109	0.00
31 S	Terphenyl-d14	0.400	0.375	6.3	107	0.00
32	Benzo(a)anthracene	0.400	0.394	1.5	110	0.00
33	Chrysene	0.400	0.405	-1.3	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.472	-18.0	141	0.00
35 I	Perylene-d12	0.400	0.400	0.0	123	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.346	13.5	109	0.02
37	Benzo(b)fluoranthene	0.400	0.354	11.5	109	0.00
38	Benzo(k)fluoranthene	0.400	0.362	9.5	112	0.00
39 C	Benzo(a)pyrene	0.400	0.366	8.5	117	0.00
40	Dibenzo(a,h)anthracene	0.400	0.345	13.8	109	0.01
41	Benzo(g,h,i)perylene	0.400	0.347	13.3	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4