

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
 Data File : BN007747.D
 Acq On : 19 Sep 2019 18:47
 Operator : HP/JU
 Sample : K4975-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT101D-20190917

Quant Time: Sep 20 11:08:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	6234	0.40	ng	0.00
7) Naphthalene-d8	10.47	136	23734	0.40	ng	0.00
13) Acenaphthene-d10	14.32	164	14309	0.40	ng	0.00
19) Phenanthrene-d10	17.07	188	32550	0.40	ng	0.00
27) Chrysene-d12	21.27	240	37050	0.40	ng	0.00
34) Perylene-d12	23.49	264	40231	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.32	112	1579	0.07	ng	0.00
5) Phenol-d6	6.87	99	1204	0.04	ng	0.00
8) Nitrobenzene-d5	8.83	82	5613	0.27	ng	0.00
11) 2-Methylnaphthalene-d10	12.06	152	12236	0.31	ng	0.00
14) 2,4,6-Tribromophenol	15.81	330	1163	0.15	ng	0.00
15) 2-Fluorobiphenyl	12.95	172	19826	0.34	ng	0.00
25) Fluoranthene-d10	19.10	212	37500	0.34	ng	0.00
29) Terphenyl-d14	19.71	244	46067	0.51	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.30	88	30747	4.425	ng	Qvalue # 92
6) bis(2-Chloroethyl)ether	7.14	93	971	0.055	ng	95
32) Bis(2-ethylhexyl)phthalate	21.20	149	3615	0.052	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
Data File : BN007747.D
Acq On : 19 Sep 2019 18:47
Operator : HP/JU
Sample : K4975-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT101D-20190917

Quant Time: Sep 20 11:08:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

