

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
 Data File : BN007741.D
 Acq On : 19 Sep 2019 15:12
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 01:48:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.446	0.406	9.0	115	0.00
3	n-Nitrosodimethylamine	0.204	0.298	-46.1#	103	0.00
4 S	2-Fluorophenol	1.377	1.029	25.3#	92	0.00
5 S	Phenol-d6	1.962	1.303	33.6#	90	0.00
6	bis(2-Chloroethyl)ether	1.133	1.013	10.6	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.350	0.296	15.4	92	0.00
9	Naphthalene	1.123	1.118	0.4	106	0.00
10	Hexachlorobutadiene	0.236	0.232	1.7	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.674	0.645	4.3	104	0.00
12	2-Methylnaphthalene	0.759	0.738	2.8	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.218	0.146	33.0#	79	0.00
15 S	2-Fluorobiphenyl	1.647	1.595	3.2	103	0.00
16	Acenaphthylene	2.121	1.818	14.3	94	-0.01
17	Acenaphthene	1.368	1.234	9.8	97	-0.01
18	Fluorene	1.753	1.587	9.5	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.243	0.241	0.8	97	0.00
21	Hexachlorobenzene	0.265	0.277	-4.5	101	0.00
22	Pentachlorophenol	0.098	0.066	32.7#	74	-0.01
23	Phenanthrene	1.246	1.252	-0.5	97	0.00
24	Anthracene	1.125	1.025	8.9	90	0.00
25 SURR	Fluoranthene-d10	1.348	1.247	7.5	90	0.00
26	Fluoranthene	1.549	1.471	5.0	92	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
28	Pvrene	1.364	1.338	1.9	91	0.00
29 S	Terphenyl-d14	0.983	0.968	1.5	91	0.00
30	Benzo(a)anthracene	1.473	1.308	11.2	83	0.00
31	Chrysene	1.436	1.465	-2.0	93	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.393	48.2#	52	0.00
33	Indeno(1,2,3-cd)pyrene	1.797	1.643	8.6	84	0.00
34 I	Perylene-d12	1.000	1.000	0.0	83	0.00
35	Benzo(b)fluoranthene	1.389	1.339	3.6	85	0.00
36	Benzo(k)fluoranthene	1.373	1.406	-2.4	89	0.00
37 C	Benzo(a)pyrene	1.305	1.182	9.4	79	0.00
38	Dibenzo(a,h)anthracene	1.335	1.274	4.6	83	0.00
39	Benzo(a,h,i)perylene	1.321	1.242	6.0	82	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
Data File : BN007741.D
Acq On : 19 Sep 2019 15:12
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 20 01:48:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		