

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
 Data File : BN007758.D
 Acq On : 20 Sep 2019 01:20
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 01:56:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.341	14.8	113	0.00
3	n-Nitrosodimethylamine	0.400	0.529	-32.3#	97	0.00
4 S	2-Fluorophenol	0.400	0.376	6.0	121	0.00
5 S	Phenol-d6	0.400	0.399	0.3	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	121	0.00
8 S	Nitrobenzene-d5	0.400	0.382	4.5	119	0.00
9	Naphthalene	0.400	0.397	0.8	121	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	122	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	127	0.00
12	2-Methylnaphthalene	0.400	0.412	-3.0	128	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	126	0.00
14 S	2,4,6-Tribromophenol	0.400	0.362	9.5	136	0.00
15 S	2-Fluorobiphenyl	0.400	0.364	9.0	123	0.00
16	Acenaphthylene	0.400	0.370	7.5	129	-0.01
17	Acenaphthene	0.400	0.355	11.3	122	-0.01
18	Fluorene	0.400	0.370	7.5	127	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.400	0.392	2.0	123	-0.01
21	Hexachlorobenzene	0.400	0.402	-0.5	125	-0.01
22	Pentachlorophenol	0.400	0.391	2.3	138	-0.01
23	Phenanthrene	0.400	0.391	2.3	122	-0.01
24	Anthracene	0.400	0.408	-2.0	130	0.00
25 SURR	Fluoranthene-d10	0.400	0.403	-0.8	125	0.00
26	Fluoranthene	0.400	0.396	1.0	123	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	113	-0.01
28	Pvrene	0.400	0.431	-7.7	126	0.00
29 S	Terphenyl-d14	0.400	0.427	-6.7	125	0.00
30	Benzo(a)anthracene	0.400	0.403	-0.8	119	0.00
31	Chrysene	0.400	0.400	0.0	116	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.470	-17.5	148	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.362	9.5	105	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	108	-0.01
35	Benzo(b)fluoranthene	0.400	0.387	3.3	111	-0.01
36	Benzo(k)fluoranthene	0.400	0.395	1.3	111	-0.01
37 C	Benzo(a)pyrene	0.400	0.390	2.5	111	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.370	7.5	104	-0.01
39	Benzo(a,h,i)perylene	0.400	0.373	6.8	105	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
Data File : BN007758.D
Acq On : 20 Sep 2019 01:20
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 20 01:56:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				