

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
 Data File : BN007758.D
 Acq On : 20 Sep 2019 01:20
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 01:56:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.446	0.380	14.8	113	0.00
3	n-Nitrosodimethylamine	0.204	0.270	-32.4#	97	0.00
4 S	2-Fluorophenol	1.377	1.294	6.0	121	0.00
5 S	Phenol-d6	1.962	1.710	12.8	123	0.00
6	bis(2-Chloroethyl)ether	1.133	1.069	5.6	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.350	0.334	4.6	119	0.00
9	Naphthalene	1.123	1.115	0.7	121	0.00
10	Hexachlorobutadiene	0.236	0.239	-1.3	122	0.00
11 SURR	2-Methylnaphthalene-d10	0.674	0.684	-1.5	127	0.00
12	2-Methylnaphthalene	0.759	0.781	-2.9	128	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
14 S	2,4,6-Tribromophenol	0.218	0.197	9.6	136	0.00
15 S	2-Fluorobiphenyl	1.647	1.497	9.1	123	0.00
16	Acenaphthylene	2.121	1.961	7.5	129	-0.01
17	Acenaphthene	1.368	1.213	11.3	122	-0.01
18	Fluorene	1.753	1.623	7.4	127	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.243	0.238	2.1	123	-0.01
21	Hexachlorobenzene	0.265	0.267	-0.8	125	-0.01
22	Pentachlorophenol	0.098	0.096	2.0	138	-0.01
23	Phenanthrene	1.246	1.218	2.2	122	-0.01
24	Anthracene	1.125	1.147	-2.0	130	0.00
25 SURR	Fluoranthene-d10	1.348	1.358	-0.7	125	0.00
26	Fluoranthene	1.549	1.535	0.9	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	113	-0.01
28	Pvrene	1.364	1.470	-7.8	126	0.00
29 S	Terphenyl-d14	0.983	1.049	-6.7	125	0.00
30	Benzo(a)anthracene	1.473	1.484	-0.7	119	0.00
31	Chrysene	1.436	1.437	-0.1	116	0.00
32	Bis(2-ethylhexyl)phthalate	0.758	0.889	-17.3	148	-0.01
33	Indeno(1,2,3-cd)pyrene	1.797	1.626	9.5	105	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	108	-0.01
35	Benzo(b)fluoranthene	1.389	1.345	3.2	111	-0.01
36	Benzo(k)fluoranthene	1.373	1.356	1.2	111	-0.01
37 C	Benzo(a)pyrene	1.305	1.274	2.4	111	-0.01
38	Dibenzo(a,h)anthracene	1.335	1.236	7.4	104	-0.01
39	Benzo(a,h,i)perylene	1.321	1.233	6.7	105	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091919\
Data File : BN007758.D
Acq On : 20 Sep 2019 01:20
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 20 01:56:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			