

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091923\
 Data File : BN027754.D
 Acq On : 19 Sep 2023 18:01
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 19:08:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 18 14:40:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	92	0.00
2	1,4-Dioxane	0.400	0.433	-8.2	89	0.00
3	n-Nitrosodimethylamine	0.400	0.438	-9.5	92	0.00
4 S	2-Fluorophenol	0.400	0.406	-1.5	90	0.00
5 S	Phenol-d6	0.400	0.404	-1.0	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.432	-8.0	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	92	0.00
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	91	0.00
9	Naphthalene	0.400	0.431	-7.7	93	0.00
10	Hexachlorobutadiene	0.400	0.436	-9.0	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.437	-9.2	95	0.00
12	2-Methylnaphthalene	0.400	0.423	-5.7	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.01
14 S	2,4,6-Tribromophenol	0.400	0.384	4.0	92	0.01
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	93	0.01
16	Acenaphthylene	0.400	0.415	-3.7	95	0.00
17	Acenaphthene	0.400	0.453	-13.2	99	0.01
18	Fluorene	0.400	0.443	-10.7	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.357	10.8	84	0.00
21	4-Bromophenyl-phenylether	0.400	0.399	0.3	94	0.01
22	Hexachlorobenzene	0.400	0.413	-3.2	95	0.00
23	Atrazine	0.400	0.373	6.8	89	0.00
24	Pentachlorophenol	0.400	0.298	25.5#	78	0.01
25	Phenanthrene	0.400	0.417	-4.2	97	0.01
26	Anthracene	0.400	0.398	0.5	94	0.00
27 SURR	Fluoranthene-d10	0.400	0.393	1.8	91	0.00
28	Fluoranthene	0.400	0.397	0.8	92	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	93	0.00
30	Pyrene	0.400	0.419	-4.7	92	0.00
31 S	Terphenyl-d14	0.400	0.392	2.0	90	0.00
32	Benzo(a)anthracene	0.400	0.394	1.5	89	0.00
33	Chrysene	0.400	0.434	-8.5	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.347	13.3	84	0.00
35 I	Perylene-d12	0.400	0.400	0.0	92	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.388	3.0	92	0.02
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	94	0.00
38	Benzo(k)fluoranthene	0.400	0.398	0.5	92	0.00
39 C	Benzo(a)pyrene	0.400	0.394	1.5	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.391	2.3	92	0.01
41	Benzo(g,h,i)perylene	0.400	0.400	0.0	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4