

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN092018\
 Data File : BN002777.D
 Acq On : 19 Sep 2018 14:06
 Operator : JU/SJ
 Sample : SSTD0.161
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.161

Quant Time: Sep 19 16:15:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN092018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 19 16:06:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	1198	0.40	ng/ul	0.00
2) Naphthalene-d8	10.46	136	5706	0.40	ng/ul	0.02
6) Acenaphthene-d10	14.30	164	3055	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.06	188	6072	0.40	ng/ul	0.01
16) Chrysene-d12	21.26	240	5189	0.40	ng/ul	0.00
20) Perylene-d12	23.48	264	5003	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.08	152	808	0.09	ng/ul	0.04
14) Fluoranthene-d10	19.10	212	1455	0.08	ng/ul	0.01

Target Compounds					Qvalue	
3) Naphthalene	10.50	128	1435	0.098	ng/ul#	77
5) 2-Methylnaphthalene	12.15	142	871	0.085	ng/ul	94
7) Acenaphthylene	14.03	152	1239	0.087	ng/ul#	86
8) Acenaphthene	14.37	153	1022	0.094	ng/ul	96
9) Fluorene	15.38	166	1060	0.085	ng/ul#	91
12) Phenanthrene	17.11	178	1625	0.090	ng/ul#	92
13) Anthracene	17.23	178	1303	0.084	ng/ul#	92
15) Fluoranthene	19.13	202	1691	0.080	ng/ul	85
17) Pyrene	19.50	202	1791	0.101	ng/ul#	88
18) Benzo(a)anthracene	21.26	228	1378	0.089	ng/ul#	87
19) Chrysene	21.30	228	1763	0.100	ng/ul	95
21) Benzo(b)fluoranthene	22.82	252	1651	0.092	ng/ul	92
22) Benzo(k)fluoranthene	22.87	252	1759	0.095	ng/ul#	89
23) Benzo(a)pyrene	23.40	252	1573	0.093	ng/ul#	52
24) Indeno(1,2,3-cd)pyrene	25.73	276	1906	0.108	ng/ul#	90
25) Dibenzo(a,h)anthracene	25.73	278	1518	0.109	ng/ul#	84
26) Benzo(g,h,i)perylene	26.40	276	1658	0.110	ng/ul#	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN092018\
Data File : BN002777.D
Acq On : 19 Sep 2018 14:06
Operator : JU/SJ
Sample : SST0.161
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.161

Quant Time: Sep 19 16:15:18 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN092018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 19 16:06:47 2018
Response via : Initial Calibration

