

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN092018\
 Data File : BN002783.D
 Acq On : 19 Sep 2018 18:12
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV67

Quant Time: Sep 19 18:50:21 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN092018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 19 17:39:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
3	Naphthalene	1.006	1.002	0.4	105	0.00
4 SURR	2-Methylnaphthalene-d10	0.577	0.576	0.2	105	0.00
5	2-Methylnaphthalene	0.652	0.659	-1.1	105	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
7	Acenaphthylene	1.689	1.651	2.2	105	0.00
8 C	Acenaphthene	1.378	1.372	0.4	107	0.00
9	Fluorene	1.471	1.476	-0.3	106	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
11	Pentachlorophenol	0.045	0.031	31.1#	73	0.01
12	Phenanthrene	1.104	1.085	1.7	103	0.00
13	Anthracene	0.959	0.947	1.3	102	0.00
14 SURR	Fluoranthene-d10	1.032	1.008	2.3	100	0.00
15 C	Fluoranthene	1.190	1.148	3.5	99	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
17	Pyrene	1.305	1.299	0.5	100	0.00
18	Benzo(a)anthracene	1.106	1.074	2.9	96	0.00
19	Chrysene	1.289	1.300	-0.9	102	0.00
20 I	Perylene-d12	1.000	1.000	0.0	98	0.00
21	Benzo(b)fluoranthene	1.322	1.285	2.8	98	0.00
22	Benzo(k)fluoranthene	1.391	1.398	-0.5	100	0.00
23 C	Benzo(a)pyrene	1.266	1.242	1.9	98	0.00
24	Indeno(1,2,3-cd)pyrene	1.526	1.503	1.5	100	0.00
25	Dibenzo(a,h)anthracene	1.222	1.209	1.1	100	0.00
26	Benzo(g,h,i)perylene	1.318	1.307	0.8	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0