

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092023\
 Data File : BN027764.D
 Acq On : 20 Sep 2023 14:37
 Operator : MA/JU
 Sample : PB155453BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB155453BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/21/2023
 Supervised By :mohammad ahmed 09/21/2023

Quant Time: Sep 20 15:28:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 20 13:45:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.008	152	2179	0.400	ng	0.00
7) Naphthalene-d8	10.831	136	6173	0.400	ng	# 0.00
13) Acenaphthene-d10	14.662	164	3985	0.400	ng	0.02
19) Phenanthrene-d10	17.418	188	11722	0.400	ng	0.02
29) Chrysene-d12	21.587	240	11418	0.400	ng	# 0.00
35) Perylene-d12	24.092	264	9936	0.400	ng	# 0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.538	112	2185	0.396	ng	0.00
5) Phenol-d6	7.163	99	2284	0.332	ng	0.00
8) Nitrobenzene-d5	9.209	82	1923	0.393	ng	0.00
11) 2-Methylnaphthalene-d10	12.430	152	3251	0.361	ng	0.02
14) 2,4,6-Tribromophenol	16.177	330	473	0.279	ng	0.04
15) 2-Fluorobiphenyl	13.304	172	5868m	0.436	ng	0.03
27) Fluoranthene-d10	19.430	212	13172	0.432	ng	0.00
31) Terphenyl-d14	20.015	244	10759	0.489	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092023\
Data File : BN027764.D
Acq On : 20 Sep 2023 14:37
Operator : MA/JU
Sample : PB155453BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB155453BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/21/2023
Supervised By :mohammad ahmed 09/21/2023

Quant Time: Sep 20 15:28:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 20 13:45:42 2023
Response via : Initial Calibration

