

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092023\
 Data File : BN027768.D
 Acq On : 20 Sep 2023 17:22
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 17:48:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 20 13:45:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	65	0.00
2	1,4-Dioxane	0.400	0.394	1.5	63	0.00
3	n-Nitrosodimethylamine	0.400	0.394	1.5	61	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	59	0.00
5 S	Phenol-d6	0.400	0.344	14.0	57	0.00
6	bis(2-Chloroethyl)ether	0.400	0.375	6.3	57	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	65	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	57	0.00
9	Naphthalene	0.400	0.388	3.0	59	0.00
10	Hexachlorobutadiene	0.400	0.391	2.3	59	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	57	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	57	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	67	0.01
14 S	2,4,6-Tribromophenol	0.400	0.359	10.3	65	0.01
15 S	2-Fluorobiphenyl	0.400	0.352	12.0	58	0.01
16	Acenaphthylene	0.400	0.377	5.8	62	0.00
17	Acenaphthene	0.400	0.401	-0.3	61	0.01
18	Fluorene	0.400	0.412	-3.0	63	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.357	10.8	69	0.00
21	4-Bromophenyl-phenylether	0.400	0.351	12.3	62	0.00
22	Hexachlorobenzene	0.400	0.373	6.8	64	0.00
23	Atrazine	0.400	0.359	10.3	65	0.00
24	Pentachlorophenol	0.400	0.353	11.8	67	0.00
25	Phenanthrene	0.400	0.389	2.8	66	0.00
26	Anthracene	0.400	0.368	8.0	65	0.00
27 SURR	Fluoranthene-d10	0.400	0.413	-3.2	70	0.00
28	Fluoranthene	0.400	0.413	-3.2	71	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	81	0.00
30	Pyrene	0.400	0.372	7.0	72	0.00
31 S	Terphenyl-d14	0.400	0.356	11.0	72	0.00
32	Benzo(a)anthracene	0.400	0.388	3.0	77	0.00
33	Chrysene	0.400	0.405	-1.3	77	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.354	11.5	75	0.00
35 I	Perylene-d12	0.400	0.400	0.0	83	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.375	6.3	80	0.00
37	Benzo(b)fluoranthene	0.400	0.379	5.3	80	0.00
38	Benzo(k)fluoranthene	0.400	0.379	5.3	80	0.00
39 C	Benzo(a)pyrene	0.400	0.374	6.5	78	0.00
40	Dibenzo(a,h)anthracene	0.400	0.375	6.3	80	0.00
41	Benzo(g,h,i)perylene	0.400	0.397	0.8	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4