

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092024\
 Data File : BN034076.D
 Acq On : 20 Sep 2024 08:23
 Operator : JU/RC
 Sample : P3957-13
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW201D-20240910

Quant Time: Sep 20 08:51:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 16:09:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.431	152	5733	0.400	ng	0.00
7) Naphthalene-d8	10.191	136	16521	0.400	ng	0.00
13) Acenaphthene-d10	14.079	164	8976	0.400	ng	0.00
19) Phenanthrene-d10	16.828	188	19353	0.400	ng	#-0.01
29) Chrysene-d12	21.051	240	13143	0.400	ng	# 0.00
35) Perylene-d12	23.178	264	11043	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.076	112	2397	0.147	ng	0.00
5) Phenol-d6	6.629	99	1549	0.082	ng	0.00
8) Nitrobenzene-d5	8.568	82	4065	0.322	ng	0.00
11) 2-Methylnaphthalene-d10	11.791	152	9036	0.370	ng	0.00
14) 2,4,6-Tribromophenol	15.587	330	1355	0.333	ng	0.00
15) 2-Fluorobiphenyl	12.690	172	13897	0.381	ng	-0.01
27) Fluoranthene-d10	18.878	212	22977	0.471	ng	0.00
31) Terphenyl-d14	19.496	244	14573	0.555	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.068	88	47387	6.611	ng	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092024\
Data File : BN034076.D
Acq On : 20 Sep 2024 08:23
Operator : JU/RC
Sample : P3957-13
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW201D-20240910

Quant Time: Sep 20 08:51:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 16:09:28 2024
Response via : Initial Calibration

