

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092024\
 Data File : BN034077.D
 Acq On : 20 Sep 2024 08:59
 Operator : JU/RC
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 09:27:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 16:09:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.397	0.8	106	0.00
3	n-Nitrosodimethylamine	0.400	0.413	-3.2	114	0.00
4 S	2-Fluorophenol	0.400	0.406	-1.5	115	0.00
5 S	Phenol-d6	0.400	0.408	-2.0	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.405	-1.3	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.392	2.0	120	0.00
9	Naphthalene	0.400	0.402	-0.5	117	0.00
10	Hexachlorobutadiene	0.400	0.405	-1.3	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	118	0.00
12	2-Methylnaphthalene	0.400	0.398	0.5	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.400	0.369	7.8	120	0.00
15 S	2-Fluorobiphenyl	0.400	0.412	-3.0	119	0.00
16	Acenaphthylene	0.400	0.391	2.3	121	0.00
17	Acenaphthene	0.400	0.403	-0.8	117	0.00
18	Fluorene	0.400	0.407	-1.7	119	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.419	-4.7	127	0.00
21	4-Bromophenyl-phenylether	0.400	0.398	0.5	119	0.00
22	Hexachlorobenzene	0.400	0.403	-0.8	116	0.00
23	Atrazine	0.400	0.368	8.0	114	0.00
24	Pentachlorophenol	0.400	0.383	4.3	129	0.00
25	Phenanthrene	0.400	0.406	-1.5	119	0.00
26	Anthracene	0.400	0.395	1.3	124	0.00
27 SURR	Fluoranthene-d10	0.400	0.387	3.3	116	0.00
28	Fluoranthene	0.400	0.389	2.8	116	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
30	Pyrene	0.400	0.459	-14.8	115	0.00
31 S	Terphenyl-d14	0.400	0.447	-11.7	113	0.00
32	Benzo(a)anthracene	0.400	0.418	-4.5	113	0.00
33	Chrysene	0.400	0.414	-3.5	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.257	35.8#	62	0.00
35 I	Perylene-d12	0.400	0.400	0.0	83	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.371	7.3	75	0.00
37	Benzo(b)fluoranthene	0.400	0.412	-3.0	92	0.00
38	Benzo(k)fluoranthene	0.400	0.438	-9.5	92	0.00
39 C	Benzo(a)pyrene	0.400	0.402	-0.5	89	0.00
40	Dibenzo(a,h)anthracene	0.400	0.366	8.5	75	0.00
41	Benzo(g,h,i)perylene	0.400	0.361	9.8	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4