

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092120\
 Data File : BN011843.D
 Acq On : 21 Sep 2020 13:33
 Operator : CG/JU
 Sample : L3988-22MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 FC-MW03SR1-20200910MS

Quant Time: Sep 21 14:02:03 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 21 12:19:13 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	1669	0.40	ng	0.00
7) Naphthalene-d8	10.491	136	6605	0.40	ng	# 0.00
13) Acenaphthene-d10	14.357	164	5783	0.40	ng	0.00
19) Phenanthrene-d10	17.102	188	10684	0.40	ng	# 0.00
29) Chrysene-d12	21.301	240	10855	0.40	ng	# 0.00
36) Perylene-d12	23.552	264	10993	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.307	112	519	0.11	ng	0.00
5) Phenol-d6	6.877	99	370	0.06	ng	0.00
8) Nitrobenzene-d5	8.856	82	1578	0.24	ng	0.00
11) 2-Methylnaphthalene-d10	12.091	152	3279	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.854	330	902	0.31	ng	0.00
15) 2-Fluorobiphenyl	12.974	172	4439	0.18	ng	0.00
27) Fluoranthene-d10	19.142	212	9454	0.28	ng	0.00
31) Terphenyl-d14	19.747	244	10282	0.41	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.261	88	478	0.19	ng	# 66
3) n-Nitrosodimethylamine	3.568	42	920	0.28	ng	# 90
6) bis(2-Chloroethyl)ether	7.135	93	1141	0.21	ng	95
9) Naphthalene	10.542	128	182091	9.79	ng	99
10) Hexachlorobutadiene	10.833	225	788	0.18	ng	# 100
12) 2-Methylnaphthalene	12.162	142	10389	0.81	ng	96
16) Acenaphthylene	14.065	152	5395	0.19	ng	# 81
17) Acenaphthene	14.420	154	4228	0.22	ng	# 89
18) Fluorene	15.410	166	8101	0.33	ng	# 97
20) 4,6-Dinitro-2-methylph...	15.486	198	504	0.28	ng	# 1
21) 4-Bromophenyl-phenylether	16.299	248	1249	0.20	ng	# 62
22) Hexachlorobenzene	16.408	284	1366	0.17	ng	98
23) Atrazine	16.579	200	957	0.17	ng	# 85
24) Pentachlorophenol	16.749	266	1580	0.50	ng	98
25) Phenanthrene	17.139	178	8163	0.24	ng	96
26) Anthracene	17.236	178	6311	0.21	ng	98
28) Fluoranthene	19.171	202	8234	0.21	ng	# 97
30) Pyrene	19.534	202	8616	0.23	ng	98
32) Benzo(a)anthracene	21.290	228	7485	0.20	ng	98
33) Chrysene	21.344	228	8029	0.21	ng	98
34) Bis(2-ethylhexyl)phtha...	21.227	149	5232	0.28	ng	97
35) Indeno(1,2,3-cd)pyrene	25.804	276	9615	0.20	ng	99
37) Benzo(b)fluoranthene	22.881	252	8559	0.21	ng	# 79
38) Benzo(k)fluoranthene	22.922	252	8524	0.20	ng	# 82
39) Benzo(a)pyrene	23.453	252	7130	0.20	ng	# 71
40) Dibenzo(a,h)anthracene	25.822	278	7682	0.19	ng	# 86
41) Benzo(g,h,i)perylene	26.496	276	8272	0.19	ng	# 93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092120\
Data File : BN011843.D
Acq On : 21 Sep 2020 13:33
Operator : CG/JU
Sample : L3988-22MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS

Quant Time: Sep 21 14:02:03 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 21 12:19:13 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.715 min Scan# 588
 Delta R.T. 0.000 min
 Lab File: BN011843.D
 Acq: 21 Sep 2020 13:33

Instrument :
 BNA_N
 ClientSampleId :
 FC-MW03SR1-20200910MS

Tgt Ion: 152 Resp: 1669




Ion	Ratio	Lower	Upper
152	100		
150	151.2	118.3	177.5
115	68.1	51.3	76.9




#2
 1,4-Dioxane
 Concen: 0.19 ng
 RT: 3.261 min Scan# 35
 Delta R.T. -0.008 min
 Lab File: BN011843.D
 Acq: 21 Sep 2020 13:33

Tgt Ion: 88 Resp: 478

Ion	Ratio	Lower	Upper
88	100		
58	81.8	63.3	94.9
43	99.0	33.0	49.4









#3
n-Nitrosodimethylamine
Concen: 0.28 ng
RT: 3.568 min Scan# 73
Delta R.T. -0.008 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion: 42 Resp: 920
Ion Ratio Lower Upper
42 100
74 71.7 64.4 96.6
44 13.8 6.7 10.1#



#4
2-Fluorophenol
Concen: 0.11 ng
RT: 5.307 min Scan# 289
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion: 112 Resp: 519
Ion Ratio Lower Upper
112 100
64 57.2 49.0 73.6
63 34.7 31.8 47.6









#5
Phenol-d6
Concen: 0.06 ng
RT: 6.877 min Scan# 484
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion: 99 Resp: 370
Ion Ratio Lower Upper
99 100
42 0.0 23.4 35.0#
71 86.2 38.1 57.1#



#6
bis(2-Chloroethyl)ether
Concen: 0.21 ng
RT: 7.135 min Scan# 516
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion: 93 Resp: 1141
Ion Ratio Lower Upper
93 100
63 78.7 61.1 91.7
95 26.4 25.5 38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.491 min Scan# 834
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:	136	Resp:	6605
Ion	Ratio	Lower	Upper
136	100		
137	12.4	9.5	14.3
54	11.3	5.6	8.4#
68	8.8	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.24 ng
RT: 8.856 min Scan# 705
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:	82	Resp:	1578
Ion	Ratio	Lower	Upper
82	100		
128	40.7	34.2	51.2
54	58.4	43.1	64.7









#9
Naphthalene
Concen: 9.79 ng
RT: 10.542 min Scan# 838
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:128 Resp: 182091
Ion Ratio Lower Upper
128 100
129 11.5 9.7 14.5
127 13.3 11.0 16.4



#10
Hexachlorobutadiene
Concen: 0.18 ng
RT: 10.833 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:225 Resp: 788
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.7 51.9 77.9









#11
2-Methylnaphthalene-d10
Concen: 0.28 ng
RT: 12.091 min Scan# 1044
Delta R.T. 0.005 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:152 Resp: 3279
Ion Ratio Lower Upper
152 100
151 18.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.81 ng
RT: 12.162 min Scan# 1060
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:142 Resp: 10389
Ion Ratio Lower Upper
142 100
141 90.4 70.0 105.0
115 48.3 35.8 53.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.357 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:	164	Resp:	5783
Ion	Ratio	Lower	Upper
164	100		
162	100.6	83.6	125.4
160	60.2	41.2	61.8



#14
2,4,6-Tribromophenol
Concen: 0.31 ng
RT: 15.854 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:	330	Resp:	902
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	0.0	33.7	50.5#









#15
2-Fluorobiphenyl
Concen: 0.18 ng
RT: 12.974 min Scan# 1174
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:	172	Resp:	4439
Ion Ratio	Lower	Upper	
172	100		
171	36.1	27.4	41.2
170	24.6	17.8	26.6



#16
Acenaphthylene
Concen: 0.19 ng
RT: 14.065 min Scan# 1260
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:	152	Resp:	5395
Ion Ratio	Lower	Upper	
152	100		
151	21.1	15.9	23.9
153	31.0	10.5	15.7#









#17
Acenaphthene
Concen: 0.22 ng
RT: 14.420 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:154 Resp: 4228
Ion Ratio Lower Upper
154 100
153 113.3 83.4 125.2
152 65.1 42.8 64.2#



#18
Fluorene
Concen: 0.33 ng
RT: 15.410 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:166 Resp: 8101
Ion Ratio Lower Upper
166 100
165 97.9 78.8 118.2
167 21.4 11.2 16.8#









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.102 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:188 Resp: 10684
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 7.7 11.5#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.28 ng
RT: 15.486 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:198 Resp: 504
Ion Ratio Lower Upper
198 100
51 247.1 45.0 67.4#
105 427.3 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.20 ng
RT: 16.299 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:248 Resp: 1249
Ion Ratio Lower Upper
248 100
250 96.5 82.7 124.1
141 113.9 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.17 ng
RT: 16.408 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:284 Resp: 1366
Ion Ratio Lower Upper
284 100
142 42.3 32.4 48.6
249 33.2 26.8 40.2





#23

Atrazine

Concen: 0.17 ng

RT: 16.579 min Scan# 1460

Delta R.T. 0.012 min

Lab File: BN011843.D

Acq: 21 Sep 2020 13:33





Tgt Ion:	200	Resp:	957
Ion Ratio	Lower	Upper	
200	100		
173	44.0	18.7	28.1#
215	52.3	40.2	60.4

Instrument :

BNA_N

ClientSampleId :

FC-MW03SR1-20200910MS



#24

Pentachlorophenol

Concen: 0.50 ng




RT: 16.749 min Scan# 1474


Delta R.T. 0.000 min

Lab File: BN011843.D

Acq: 21 Sep 2020 13:33

Tgt Ion:	266	Resp:	1580
Ion Ratio	Lower	Upper	
266	100		
264	63.9	51.2	76.8
268	63.1	52.4	78.6









#25
Phenanthrene
Concen: 0.24 ng
RT: 17.139 min Scan# 1506
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:178 Resp: 8163
Ion Ratio Lower Upper
178 100
176 20.5 15.1 22.7
179 17.2 12.3 18.5



#26
Anthracene
Concen: 0.21 ng
RT: 17.236 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:178 Resp: 6311
Ion Ratio Lower Upper
178 100
176 18.6 14.6 21.8
179 16.8 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.28 ng
RT: 19.142 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:212 Resp: 9454
Ion Ratio Lower Upper
212 100
106 10.0 6.4 9.6#
104 6.0 3.9 5.9#



#28
Fluoranthene
Concen: 0.21 ng
RT: 19.171 min Scan# 1811
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:202 Resp: 8234
Ion Ratio Lower Upper
202 100
101 9.4 5.4 8.2#
203 17.8 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.301 min Scan# 2101
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:240 Resp: 10855
Ion Ratio Lower Upper
240 100
120 12.5 3.5 5.3#
236 28.1 20.5 30.7



#30
Pyrene
Concen: 0.23 ng
RT: 19.534 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:202 Resp: 8616
Ion Ratio Lower Upper
202 100
200 21.1 16.5 24.7
203 18.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.41 ng
RT: 19.747 min Scan# 1927
Delta R.T. 0.005 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:244	Resp:	10282
Ion Ratio	Lower	Upper
244	100	
212	9.1	7.3 10.9
122	9.9	7.8 11.6



#32
Benzo(a)anthracene
Concen: 0.20 ng
RT: 21.290 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:228	Resp:	7485
Ion Ratio	Lower	Upper
228	100	
226	27.5	21.0 31.4
229	20.4	16.0 24.0





#33

Chrysene

Concen: 0.21 ng

RT: 21.344 min Scan# 2105

Delta R.T. 0.000 min

Lab File: BN011843.D

Acq: 21 Sep 2020 13:33

Instrument :





BNA_N

ClientSampleId :

FC-MW03SR1-20200910MS

Tgt Ion:228 Resp: 8029

Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.3	34.9
229	20.8	15.8	23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.28 ng

RT: 21.227 min Scan# 2094




Delta R.T. 0.000 min


Lab File: BN011843.D

Acq: 21 Sep 2020 13:33

Tgt Ion:149 Resp: 5232

Ion	Ratio	Lower	Upper
149	100		
167	27.9	23.6	35.4
279	4.7	4.0	6.0









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.20 ng
RT: 25.804 min Scan# 3198
Delta R.T. -0.003 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33




Instrument :
BNA_N
ClientSampleId :
FC-MW03SR1-20200910MS


Tgt Ion:276 Resp: 9615
Ion Ratio Lower Upper
276 100
138 16.2 13.0 19.4
277 24.1 20.0 30.0



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.552 min Scan# 2540
Delta R.T. 0.000 min
Lab File: BN011843.D
Acq: 21 Sep 2020 13:33

Tgt Ion:264 Resp: 10993
Ion Ratio Lower Upper
264 100
260 25.0 19.6 29.4
265 70.3 38.6 58.0#





#37

Benzo(b)fluoranthene

Concen: 0.21 ng

RT: 22.881 min Scan# 2344

Delta R.T. 0.003 min

Lab File: BN011843.D

Acq: 21 Sep 2020 13:33

Tgt Ion:252 Resp: 8559





Ion	Ratio	Lower	Upper
252	100		
253	32.5	19.1	28.7#
125	18.7	6.2	9.2#

Instrument :

BNA_N

ClientSampleId :

FC-MW03SR1-20200910MS



#38

Benzo(k)fluoranthene

Concen: 0.20 ng

RT: 22.922 min Scan# 2356




Delta R.T. 0.000 min


Lab File: BN011843.D

Acq: 21 Sep 2020 13:33

Tgt Ion:252 Resp: 8524

Ion	Ratio	Lower	Upper
252	100		
253	32.0	19.3	28.9#
125	16.2	6.0	9.0#





#39

Benzo(a)pyrene

Concen: 0.20 ng

RT: 23.453 min Scan# 2511

Delta R.T. 0.000 min

Lab File: BN011843.D

Acq: 21 Sep 2020 13:33





Tgt Ion:252	Resp:	7130
Ion Ratio	Lower	Upper
252	100	
253	36.2	19.8 29.6#
125	25.7	7.3 10.9#

Instrument :

BNA_N

ClientSampleId :

FC-MW03SR1-20200910MS



#40

Dibenzo(a,h)anthracene

Concen: 0.19 ng




RT: 25.822 min Scan# 3203


Delta R.T. 0.000 min

Lab File: BN011843.D

Acq: 21 Sep 2020 13:33


Tgt Ion:278	Resp:	7682
Ion Ratio	Lower	Upper
278	100	
139	16.2	8.6 13.0#
279	32.0	19.8 29.8#





#41
 Benzo(g,h,i)perylene
 Concen: 0.19 ng
 RT: 26.496 min Scan# 3400
 Delta R.T. 0.004 min
 Lab File: BN011843.D
 Acq: 21 Sep 2020 13:33

Instrument :
 BNA_N
 ClientSampleId :
 FC-MW03SR1-20200910MS



Tgt Ion:	276	Resp:	8272
Ion Ratio	Lower	Upper	
276	100		
277	26.5	19.4	29.0
138	18.8	11.3	16.9#

