

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092124\
 Data File : BN034096.D
 Acq On : 20 Sep 2024 20:32
 Operator : JU/RC
 Sample : P4003-08
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D1-20240912

Quant Time: Sep 21 01:20:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 16:09:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.430	152	4757	0.400	ng	0.00
7) Naphthalene-d8	10.191	136	13453	0.400	ng	0.00
13) Acenaphthene-d10	14.079	164	7965	0.400	ng	0.00
19) Phenanthrene-d10	16.828	188	17794	0.400	ng	#-0.01
29) Chrysene-d12	21.051	240	13093	0.400	ng	0.00
35) Perylene-d12	23.178	264	11114	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.069	112	1668	0.124	ng	0.00
5) Phenol-d6	6.629	99	995	0.063	ng	0.00
8) Nitrobenzene-d5	8.568	82	3423	0.333	ng	0.00
11) 2-Methylnaphthalene-d10	11.788	152	7383	0.372	ng	0.00
14) 2,4,6-Tribromophenol	15.574	330	1163	0.322	ng	-0.01
15) 2-Fluorobiphenyl	12.690	172	11882	0.367	ng	-0.01
27) Fluoranthene-d10	18.878	212	19910	0.443	ng	0.00
31) Terphenyl-d14	19.496	244	12649	0.484	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.075	88	427	0.072	ng	# 55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092124\
Data File : BN034096.D
Acq On : 20 Sep 2024 20:32
Operator : JU/RC
Sample : P4003-08
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D1-20240912

Quant Time: Sep 21 01:20:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 16:09:28 2024
Response via : Initial Calibration

