

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092124\
 Data File : BN034115.D
 Acq On : 21 Sep 2024 08:42
 Operator : JU/RC
 Sample : P4116-10
 Misc :
 ALS Vial : 78 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT15811-20240916

Quant Time: Sep 22 23:57:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 16:09:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.430	152	5821	0.400	ng	0.00
7) Naphthalene-d8	10.180	136	16502	0.400	ng	#-0.02
13) Acenaphthene-d10	14.072	164	9264	0.400	ng	0.00
19) Phenanthrene-d10	16.833	188	19844	0.400	ng	0.00
29) Chrysene-d12	21.053	240	14332	0.400	ng	0.00
35) Perylene-d12	23.174	264	11775	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.069	112	2379	0.144	ng	0.00
5) Phenol-d6	6.629	99	1496	0.078	ng	0.00
8) Nitrobenzene-d5	8.568	82	3988	0.316	ng	0.00
11) 2-Methylnaphthalene-d10	11.784	152	8295	0.340	ng	0.00
14) 2,4,6-Tribromophenol	15.579	330	1218	0.290	ng	0.00
15) 2-Fluorobiphenyl	12.693	172	13390	0.355	ng	0.00
27) Fluoranthene-d10	18.876	212	19747	0.394	ng	-0.01
31) Terphenyl-d14	19.494	244	13182	0.461	ng	-0.01
Target Compounds						
24) Pentachlorophenol	16.498	266	233	0.056	ng	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092124\
Data File : BN034115.D
Acq On : 21 Sep 2024 08:42
Operator : JU/RC
Sample : P4116-10
Misc :
ALS Vial : 78 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT15811-20240916

Quant Time: Sep 22 23:57:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 16:09:28 2024
Response via : Initial Calibration

