

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092124\
 Data File : BN034137.D
 Acq On : 21 Sep 2024 23:22
 Operator : JU/RC
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 102 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 00:01:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 16:09:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	87	-0.01
2	1,4-Dioxane	0.400	0.398	0.5	86	-0.01
3	n-Nitrosodimethylamine	0.400	0.461	-15.3	104	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	88	-0.01
5 S	Phenol-d6	0.400	0.386	3.5	93	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	88	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	88	-0.02
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	97	0.00
9	Naphthalene	0.400	0.403	-0.8	93	0.00
10	Hexachlorobutadiene	0.400	0.420	-5.0	94	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	91	0.00
12	2-Methylnaphthalene	0.400	0.398	0.5	93	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.370	7.5	91	0.00
15 S	2-Fluorobiphenyl	0.400	0.420	-5.0	92	-0.02
16	Acenaphthylene	0.400	0.410	-2.5	96	0.00
17	Acenaphthene	0.400	0.418	-4.5	92	0.00
18	Fluorene	0.400	0.415	-3.7	92	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	87	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.398	0.5	90	0.00
21	4-Bromophenyl-phenylether	0.400	0.403	-0.8	92	0.00
22	Hexachlorobenzene	0.400	0.411	-2.7	91	0.00
23	Atrazine	0.400	0.385	3.8	92	0.00
24	Pentachlorophenol	0.400	0.427	-6.7	110	0.00
25	Phenanthrene	0.400	0.407	-1.7	92	0.00
26	Anthracene	0.400	0.401	-0.3	97	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	89	-0.02
28	Fluoranthene	0.400	0.398	0.5	91	-0.01
29 I	Chrysene-d12	0.400	0.400	0.0	82	0.00
30	Pyrene	0.400	0.442	-10.5	91	-0.01
31 S	Terphenyl-d14	0.400	0.430	-7.5	89	-0.01
32	Benzo(a)anthracene	0.400	0.406	-1.5	90	0.00
33	Chrysene	0.400	0.420	-5.0	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.120	70.0#	24	0.00
35 I	Perylene-d12	0.400	0.400	0.0	70	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.354	11.5	60	-0.03
37	Benzo(b)fluoranthene	0.400	0.421	-5.2	79	-0.01
38	Benzo(k)fluoranthene	0.400	0.438	-9.5	77	-0.01
39 C	Benzo(a)pyrene	0.400	0.405	-1.3	75	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.350	12.5	60	-0.03
41	Benzo(g,h,i)perylene	0.400	0.339	15.3	59	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4