

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092218\
 Data File : BN002821.D
 Acq On : 21 Sep 2018 16:58
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.472

Quant Time: Sep 21 23:10:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN092018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 21 23:09:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
3	Naphthalene	1.006	1.036	-3.0	101	0.00
4 SURR	2-Methylnaphthalene-d10	0.577	0.598	-3.6	102	0.00
5	2-Methylnaphthalene	0.652	0.688	-5.5	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
7	Acenaphthylene	1.689	1.725	-2.1	106	0.00
8 C	Acenaphthene	1.378	1.415	-2.7	106	0.00
9	Fluorene	1.471	1.553	-5.6	107	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
11	Pentachlorophenol	0.045	0.052	-15.6	119	0.00
12	Phenanthrene	1.104	1.160	-5.1	106	0.00
13	Anthracene	0.959	1.030	-7.4	107	0.00
14 SURR	Fluoranthene-d10	1.032	1.105	-7.1	105	0.00
15 C	Fluoranthene	1.190	1.259	-5.8	104	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
17	Pyrene	1.305	1.423	-9.0	104	0.00
18	Benzo(a)anthracene	1.106	1.197	-8.2	102	0.00
19	Chrysene	1.289	1.369	-6.2	102	0.00
20 I	Perylene-d12	1.000	1.000	0.0	95	0.00
21	Benzo(b)fluoranthene	1.322	1.381	-4.5	102	0.00
22	Benzo(k)fluoranthene	1.391	1.462	-5.1	101	0.00
23 C	Benzo(a)pyrene	1.266	1.322	-4.4	101	0.00
24	Indeno(1,2,3-cd)pyrene	1.526	1.557	-2.0	100	0.00
25	Dibenzo(a,h)anthracene	1.222	1.244	-1.8	100	0.00
26	Benzo(g,h,i)perylene	1.318	1.354	-2.7	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0