


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092218\
 Data File : BN002808.D
 Acq On : 21 Sep 2018 08:17
 Operator : JU/SJ
 Sample : DFTPP24
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP24

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SOM-EPA-SIM-BN092018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 24 16:43:54 2018



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.2	95333	PASS
68	69	0.00	2	1.6	1671	PASS
69	198	0.00	100	36.9	102997	PASS
70	69	0.00	2	0.5	537	PASS
127	198	10	80	45.0	125664	PASS
197	198	0.00	2	0.2	569	PASS
198	198	100	100	100.0	279146	PASS
199	198	5	9	6.9	19264	PASS
275	198	10	60	27.5	76885	PASS
365	198	1	100	3.2	8857	PASS
441	198	0.01	100	14.1	39242	PASS
442	198	50	100	91.3	254997	PASS
443	442	15	24	18.8	47829	PASS

m/z	Abundance
37.00	425.0
38.10	964.0
39.00	7228.0
39.90	695.0
44.00	2526.0
49.00	962.0
50.00	25304.0
51.00	100296.0
52.10	5184.0
54.90	459.0
56.00	3126.0
57.00	7624.0
57.90	318.0
61.00	1330.0
62.00	1742.0
63.00	4071.0
64.00	632.0
65.00	2172.0
68.10	1974.0
69.00	106600.0
70.10	584.0
72.90	1116.0
74.00	9218.0
75.00	16408.0
76.00	5373.0
77.00	119456.0
78.10	7761.0
79.00	7402.0
80.00	5345.0
81.00	7903.0
82.00	2109.0
83.00	1733.0
84.90	1301.0
86.00	2296.0
87.00	999.0
87.90	442.0
90.90	1594.0
92.00	1989.0
93.00	11770.0
94.00	863.0
96.00	539.0
96.80	325.0
98.00	9273.0
99.00	7772.0
99.90	831.0
101.00	4801.0
101.90	324.0
103.00	1532.0
104.00	3109.0
105.00	2550.0
106.00	845.0
107.00	33704.0
108.00	5601.0
110.00	64776.0
111.00	9714.0
112.00	1445.0
113.10	341.0
116.00	1835.0
117.00	23072.0
118.00	1735.0
120.00	440.0
122.00	1904.0
123.00	3366.0
123.90	1682.0
124.90	1499.0
127.00	122976.0
128.00	9302.0
129.00	45448.0
130.00	3529.0
131.10	768.0
131.90	615.0
132.90	324.0
133.90	1301.0
135.00	3656.0
135.90	1318.0
136.90	2123.0
137.80	660.0
139.90	767.0
140.90	5712.0
142.00	1878.0
142.90	1845.0
144.10	340.0
145.90	887.0
147.00	2965.0
147.90	5941.0
148.90	1610.0
149.90	588.0
151.00	832.0
152.00	507.0
153.00	2050.0
154.00	1437.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

155.00	3168.0
156.00	4748.0
156.90	897.0
157.10	884.0
157.90	1319.0
158.90	699.0
159.90	1444.0
161.00	2940.0
161.90	654.0
163.80	352.0
164.90	2105.0
166.00	1942.0
167.00	9941.0
168.00	5943.0
168.90	994.0
170.00	335.0
170.90	787.0
171.90	915.0
173.10	1515.0
174.00	2334.0
175.00	4987.0
176.00	1494.0
176.90	2446.0
177.90	869.0
179.00	8678.0
180.00	6501.0
181.00	2973.0
184.00	840.0
185.00	4346.0
186.00	33896.0
186.90	9421.0
188.00	1110.0
188.90	2017.0
191.00	1140.0
192.00	2584.0
193.00	3522.0
194.00	897.0
195.10	403.0
196.00	7195.0
196.80	1708.0
197.90	255360.0
199.00	17936.0
199.90	1247.0
201.50	1204.0
202.90	1589.0
204.00	8677.0
205.00	14801.0
206.00	65512.0
207.00	9582.0
208.00	2328.0
208.90	741.0
210.90	2348.0
211.80	673.0
214.90	529.0
215.90	1442.0
216.90	15776.0
217.90	1688.0
221.00	10152.0
221.80	2129.0
223.00	3960.0
224.00	35640.0
225.00	7672.0
226.10	1119.0
226.90	12890.0
227.90	1810.0
228.90	2618.0
230.00	444.0
231.00	1285.0
233.90	752.0
235.00	983.0
235.90	652.0
237.00	1677.0
239.00	598.0
239.90	352.0
240.90	765.0
242.00	1757.0
243.00	2147.0
244.00	27296.0
245.00	3082.0
245.90	5299.0
246.90	1119.0
248.90	938.0
253.00	618.0
255.00	139584.0
256.00	20696.0
256.90	1978.0
258.00	6838.0
259.00	1617.0
264.00	317.0
265.00	2827.0
266.00	394.0
270.90	479.0
271.90	453.0
273.00	4250.0
274.00	11740.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

275.00	63752.0
276.00	9757.0
277.00	4897.0
278.00	903.0
282.90	719.0
283.80	340.0
285.10	857.0
286.00	338.0
293.10	965.0
293.90	420.0
295.00	315.0
295.90	14960.0
297.00	2343.0
303.00	1931.0
303.90	471.0
307.90	376.0
313.90	600.0
315.00	1530.0
315.90	984.0
320.90	479.0
323.00	5208.0
324.00	735.0
326.90	1136.0
327.80	519.0
331.90	399.0
332.90	539.0
334.00	3528.0
335.00	888.0
340.90	667.0
345.90	871.0
352.00	1789.0
353.00	1139.0
354.00	1742.0
355.00	501.0
365.00	6482.0
365.90	1112.0
370.90	496.0
372.10	2703.0
373.10	919.0
382.80	613.0
383.00	573.0
389.90	431.0
402.00	1154.0
403.00	1963.0
403.90	970.0
420.90	1341.0
422.00	1756.0
423.00	10660.0
424.00	3040.0
424.80	627.0
441.00	27976.0
442.00	181952.0
443.00	33544.0
444.00	4313.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

m/z	Abundance
37.00	420.0
38.00	1116.0
39.00	8280.0
40.00	800.0
41.00	345.0
44.00	2406.0
45.00	320.0
49.10	763.0
50.00	27776.0
51.00	114056.0
52.00	5843.0
54.90	618.0
55.90	3933.0
57.00	8650.0
57.90	430.0
61.00	1687.0
61.90	1620.0
63.10	4900.0
63.90	556.0
65.10	2425.0
68.00	1840.0
69.00	122632.0
70.00	584.0
72.90	987.0
74.00	11578.0
75.00	19304.0
76.00	6296.0
77.00	136448.0
78.00	9139.0
79.00	8758.0
80.00	6460.0
81.00	9654.0
82.00	2215.0
82.90	2463.0
85.00	2012.0
86.00	2297.0
87.00	1163.0
87.90	394.0
90.90	2374.0
92.00	2280.0
93.00	13934.0
94.00	1199.0
96.00	693.0
96.70	319.0
98.00	9933.0
99.00	8697.0
99.90	801.0
101.00	5638.0
101.90	318.0
102.90	2000.0
104.00	3317.0
105.00	2717.0
106.10	1086.0
107.00	41104.0
108.00	6508.0
109.00	1180.0
110.00	78184.0
111.00	10887.0
112.00	1226.0
112.90	449.0
116.00	2133.0
117.00	27304.0
118.00	1915.0
119.90	589.0
121.90	2315.0
123.00	4296.0
124.00	1790.0
125.00	1796.0
127.00	149824.0
128.00	10968.0
129.00	56904.0
130.00	4362.0
131.00	944.0
132.20	561.0
132.70	498.0
133.90	1828.0
135.00	4110.0
135.90	1592.0
137.00	2719.0
138.10	512.0
140.00	705.0
141.00	7382.0
142.00	2483.0
142.90	1831.0
144.00	411.0
144.90	473.0
145.80	1237.0
147.00	3584.0
148.00	7940.0
149.00	1537.0
149.90	403.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

151.20	1116.0
152.00	463.0
152.90	2168.0
154.00	1775.0
155.00	3575.0
156.00	6504.0
157.00	1353.0
158.00	1492.0
159.00	1146.0
160.00	2306.0
161.00	3770.0
162.00	940.0
162.90	381.0
164.00	441.0
165.00	2261.0
166.00	2021.0
167.00	14340.0
168.00	7780.0
168.90	1398.0
169.90	424.0
170.90	764.0
172.00	1174.0
172.90	2085.0
174.00	3210.0
175.00	5730.0
176.00	1731.0
176.90	2499.0
178.10	899.0
178.90	11046.0
180.00	8133.0
180.90	3472.0
181.90	554.0
183.00	358.0
183.90	960.0
184.90	5552.0
186.00	43352.0
187.00	12269.0
188.00	1423.0
188.90	2911.0
190.00	556.0
191.00	1639.0
192.00	3709.0
192.90	3942.0
194.00	745.0
194.90	434.0
196.00	9792.0
197.90	337856.0
199.00	23328.0
200.00	1882.0
201.40	1713.0
203.00	2383.0
204.00	11456.0
205.00	19096.0
206.00	84016.0
207.00	12106.0
208.00	2799.0
209.00	965.0
210.10	1104.0
211.00	2988.0
214.90	1018.0
216.00	2001.0
216.90	20808.0
217.90	2919.0
218.80	362.0
219.90	590.0
221.00	14284.0
223.00	4969.0
224.00	46320.0
225.00	11360.0
226.00	1374.0
227.00	18328.0
227.90	3026.0
229.00	3957.0
230.00	636.0
231.00	1734.0
231.90	361.0
233.00	377.0
233.90	975.0
234.90	1531.0
235.90	1009.0
237.00	1587.0
238.00	434.0
239.00	655.0
240.00	617.0
240.90	1309.0
242.00	2655.0
243.00	2652.0
244.00	36352.0
245.00	4998.0
245.90	6445.0
246.90	1263.0
248.90	1536.0
250.90	420.0
251.80	330.0
253.00	837.0

Instrument :

BNA_N

ClientSampleId :

DFTPP24

255.00	196480.0
256.00	29160.0
257.00	2037.0
258.00	10550.0
259.00	1651.0
264.90	4109.0
265.90	745.0
271.10	354.0
271.90	569.0
273.00	6119.0
274.00	14563.0
275.00	93744.0
276.00	12861.0
277.00	7595.0
278.00	1308.0
282.00	322.0
283.00	763.0
283.90	488.0
285.00	1401.0
289.00	313.0
291.90	394.0
293.00	1950.0
293.90	447.0
296.00	22528.0
297.00	2946.0
300.90	339.0
301.80	707.0
303.00	2839.0
304.00	830.0
308.00	451.0
309.90	403.0
314.00	939.0
315.00	2505.0
316.10	1322.0
317.00	427.0
321.00	645.0
321.90	355.0
323.00	8586.0
324.00	1423.0
326.90	1368.0
328.00	823.0
331.90	523.0
332.90	686.0
334.00	5240.0
334.90	1141.0
340.90	1040.0
345.90	1555.0
346.90	318.0
352.00	2479.0
353.00	1985.0
354.10	2648.0
355.00	502.0
358.90	328.0
364.90	10822.0
365.90	1300.0
370.90	691.0
372.00	4545.0
372.90	1030.0
382.90	1514.0
384.00	433.0
389.90	656.0
390.90	415.0
392.00	421.0
400.90	355.0
401.90	1872.0
403.00	2800.0
404.00	930.0
420.90	2179.0
422.00	2319.0
423.00	16293.0
424.00	4559.0
425.00	594.0
441.00	46584.0
442.00	299456.0
443.00	55848.0
444.00	4925.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

m/z	Abundance
36.90	309.0
38.10	734.0
39.00	5403.0
40.00	478.0
44.00	2833.0
49.00	339.0
50.00	17320.0
51.00	71648.0
52.00	3638.0
55.00	330.0
56.00	2106.0
57.00	4992.0
58.10	340.0
61.00	1089.0
62.00	1226.0
63.10	3015.0
63.90	533.0
64.90	1427.0
68.10	1201.0
69.00	79760.0
69.90	443.0
73.00	973.0
74.00	6739.0
75.00	11831.0
76.00	3051.0
77.00	87392.0
78.00	6122.0
79.00	5336.0
80.00	3667.0
80.90	5919.0
81.90	1414.0
83.00	1879.0
85.00	1246.0
86.00	1613.0
87.00	729.0
91.00	1364.0
91.90	1358.0
93.00	8982.0
94.00	602.0
95.90	597.0
98.00	6853.0
99.00	5633.0
100.00	710.0
101.00	3447.0
103.00	1391.0
104.00	2251.0
105.00	1930.0
106.00	638.0
107.00	26048.0
108.00	4923.0
109.00	634.0
110.00	51336.0
111.00	8010.0
111.90	912.0
116.00	1431.0
117.00	18912.0
118.00	1280.0
120.00	418.0
122.00	2109.0
122.90	2983.0
124.00	1149.0
124.90	1116.0
127.00	104192.0
128.00	7262.0
129.00	36928.0
130.00	2989.0
131.00	557.0
132.00	349.0
134.00	990.0
135.00	3223.0
136.10	1276.0
137.00	1353.0
137.90	442.0
140.10	426.0
141.00	4726.0
142.00	1657.0
142.80	1204.0
145.90	768.0
147.00	2979.0
147.90	4945.0
148.90	1165.0
150.10	349.0
151.00	1015.0
151.70	510.0
152.90	1735.0
154.00	1377.0
155.00	2833.0
156.00	4577.0
157.00	900.0
157.80	769.0
158.90	810.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

160.00	1674.0
161.00	2353.0
161.90	918.0
164.90	1970.0
166.00	1296.0
167.00	9724.0
167.90	4760.0
169.00	887.0
171.00	710.0
172.00	804.0
172.90	1404.0
174.00	2107.0
174.90	4522.0
176.00	1263.0
177.00	2183.0
178.00	755.0
178.90	7947.0
180.00	5150.0
181.00	2503.0
182.10	391.0
183.90	593.0
185.00	3518.0
186.00	32064.0
187.00	8661.0
188.00	739.0
188.90	1833.0
191.00	1098.0
192.00	2292.0
193.00	2875.0
193.90	700.0
195.10	472.0
195.90	6829.0
197.90	244224.0
198.90	16528.0
199.90	1262.0
201.50	1020.0
202.90	1513.0
204.00	8619.0
205.00	13411.0
206.00	59336.0
207.00	8868.0
208.00	2238.0
209.00	982.0
210.00	830.0
210.90	2819.0
212.10	395.0
214.90	712.0
216.00	1179.0
217.00	14967.0
218.00	1908.0
221.00	11143.0
223.00	3596.0
224.00	34792.0
225.00	8454.0
226.00	1142.0
227.00	14008.0
228.00	1722.0
229.00	2876.0
231.00	1400.0
232.90	318.0
234.00	1147.0
234.90	1190.0
236.00	701.0
237.00	1500.0
239.00	515.0
240.00	462.0
240.90	706.0
241.90	1962.0
243.10	1883.0
244.00	28768.0
245.00	3828.0
245.90	4755.0
247.00	1041.0
249.00	1219.0
252.90	682.0
254.00	1374.0
255.00	149888.0
256.00	22296.0
257.00	1561.0
257.90	6847.0
259.00	1619.0
264.90	3089.0
265.90	507.0
273.00	4015.0
274.00	12188.0
275.00	73160.0
276.00	9480.0
277.00	5346.0
277.90	810.0
281.90	325.0
283.00	837.0
284.00	438.0
285.00	1107.0
288.90	314.0
292.90	1363.0

Instrument :

BNA_N

ClientSampleId :

DFTPP24

296.00	17760.0
296.90	2602.0
301.00	311.0
301.90	377.0
302.90	2043.0
304.00	696.0
308.00	376.0
314.00	1129.0
315.00	2042.0
316.10	922.0
316.90	331.0
320.90	482.0
321.80	405.0
323.00	7041.0
324.00	1175.0
326.90	1344.0
328.00	838.0
332.00	594.0
332.80	766.0
334.00	4192.0
335.00	1237.0
341.00	903.0
346.00	1434.0
352.00	2275.0
353.00	1642.0
354.00	2510.0
355.00	388.0
365.00	9269.0
365.90	1255.0
371.00	671.0
372.00	4215.0
373.00	687.0
383.00	1436.0
383.90	311.0
389.90	535.0
391.00	301.0
391.90	348.0
402.10	1792.0
403.00	2682.0
404.00	1166.0
421.00	2066.0
422.10	2181.0
423.00	16536.0
424.00	4091.0
425.00	423.0
441.00	43168.0
442.00	283584.0
443.00	54096.0
444.00	4642.0
445.00	319.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24

Scan 2456 (17.433 min): BN002808.D
DFTPP24

m/z	Abundance
39.90	350.0
43.90	2312.0
72.90	333.0
207.00	1170.0

Instrument :
BNA_N
ClientSampleId :
DFTPP24