

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN092319\
 Data File : BN007803.D
 Acq On : 23 Sep 2019 16:08
 Operator : HP/JU
 Sample : PB123286BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK86

Quant Time: Sep 23 17:02:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 23 15:18:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	429849	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	1767589	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	1173580	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	2516620	20.00	ng/ul	0.00
77) Chrysene-d12	21.26	240	2680765	20.00	ng/ul	0.00
85) Perylene-d12	23.47	264	2964944	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.25	96	85949	7.60	ng/uL	0.00
5) Phenol-d5	6.86	99	1559997	38.11	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.02	67	945949	40.85	ng/ul	0.00
9) 2-Chlorophenol-d4	7.22	132	1166427	40.74	ng/ul	0.00
13) 4-Methylphenol-d8	8.38	113	1184419	37.97	ng/ul	0.00
19) Nitrobenzene-d5	8.83	128	552461	42.83	ng/ul	0.00
22) 2-Nitrophenol-d4	9.55	143	558379	45.60	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.07	165	1140297	39.66	ng/ul	0.00
29) 4-Chloroaniline-d4	10.59	131	1609348	43.15	ng/ul	0.00
43) Dimethylphthalate-d6	13.73	166	3836224	41.08	ng/ul	0.00
46) Acenaphthylene-d8	14.00	160	4231353	39.56	ng/ul	0.00
51) 4-Nitrophenol-d4	14.51	143	558752	33.64	ng/ul	0.00
57) Fluorene-d10	15.31	176	3203034	39.53	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	405464	29.85	ng/ul	0.00
70) Anthracene-d10	17.16	188	5172062	41.29	ng/ul	0.00
78) Pyrene-d10	19.46	212	6138720	42.51	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.34	264	6550755	42.46	ng/ul	0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN092319\
Data File : BN007803.D
Acq On : 23 Sep 2019 16:08
Operator : HP/JU
Sample : PB123286BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK86

Quant Time: Sep 23 17:02:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091619MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 23 15:18:07 2019
Response via : Initial Calibration

