


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092419\
 Data File : BN007831.D
 Acq On : 24 Sep 2019 11:19
 Operator : JU
 Sample : DFTPP52
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP52

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN091619MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Sep 25 08:12:46 2019



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.4	227960	PASS
68	69	0.00	2	1.9	6173	PASS
69	198	0.00	100	36.1	323541	PASS
70	69	0.00	2	0.6	2025	PASS
127	198	10	80	39.8	357248	PASS
197	198	0.00	2	0.7	6070	PASS
198	198	100	100	100.0	896832	PASS
199	198	5	9	6.8	61096	PASS
275	198	10	60	27.3	244821	PASS
365	198	1	100	3.0	26704	PASS
441	198	0.01	100	10.5	94157	PASS
442	198	50	100	67.2	602304	PASS
443	442	15	24	19.6	117909	PASS

m/z	Abundance
38.90	491.0
39.90	2550.0
40.90	574.0
41.90	348.0
42.90	662.0
44.00	12657.0
48.90	1064.0
50.90	754.0
53.00	314.0
55.10	516.0
72.90	1021.0
77.80	324.0
80.90	365.0
83.90	1318.0
85.90	1085.0
91.00	375.0
95.90	763.0
133.00	417.0
135.10	300.0
146.90	416.0
190.80	799.0
193.10	329.0
207.00	5097.0
208.00	1079.0
208.90	688.0
280.90	779.0

Instrument :

BNA_N

ClientSampleId :

DFTPP52

m/z	Abundance
36.00	637.0
37.00	1100.0
38.00	3331.0
39.00	18976.0
40.00	3811.0
41.00	895.0
43.00	662.0
44.00	10920.0
44.90	859.0
49.00	2875.0
50.00	64464.0
51.00	257408.0
52.00	13717.0
53.00	973.0
55.00	1793.0
56.00	8168.0
57.00	18360.0
58.00	892.0
59.00	493.0
59.90	384.0
61.00	3735.0
62.00	4330.0
63.00	13905.0
64.10	1869.0
65.10	6756.0
66.10	676.0
66.90	660.0
68.00	6887.0
69.00	365504.0
70.00	2314.0
73.00	3392.0
74.00	37760.0
75.00	57624.0
76.00	18952.0
77.00	410304.0
78.00	26616.0
79.00	26328.0
80.00	20328.0
81.00	27488.0
82.00	6964.0
83.00	6992.0
84.00	2060.0
85.00	4592.0
86.00	8188.0
87.00	3606.0
88.00	1738.0
88.90	662.0
91.00	6399.0
92.00	7189.0
93.00	42176.0
93.90	2847.0
95.00	683.0
96.00	2734.0
96.90	925.0
98.00	32656.0
99.00	31096.0
100.00	2748.0
101.00	16015.0
101.90	883.0
103.00	5605.0
104.00	9676.0
105.00	8576.0
106.00	2749.0
107.00	109432.0
108.00	17344.0
109.10	3933.0
110.00	213504.0
111.00	32424.0
112.00	3554.0
113.10	1113.0
115.00	504.0
116.00	6499.0
117.00	84440.0
118.00	7203.0
119.00	1345.0
120.00	1283.0
120.90	727.0
122.00	6619.0
123.00	11105.0
123.90	5436.0
124.90	4869.0
127.00	370112.0
128.00	29000.0
129.00	138432.0
130.00	13255.0
131.00	1907.0
132.00	1345.0
133.10	1053.0
134.00	4004.0
135.00	11251.0
135.90	5131.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

137.00	7787.0
137.90	2024.0
139.00	1046.0
140.10	1873.0
141.00	18160.0
142.00	6726.0
143.00	3489.0
143.90	892.0
145.00	1036.0
146.00	3130.0
147.00	10053.0
148.00	21800.0
149.00	4935.0
150.00	1321.0
151.20	2172.0
152.00	1716.0
153.00	6486.0
154.00	4338.0
155.00	11308.0
156.00	15230.0
157.00	3296.0
157.90	3739.0
158.90	2724.0
159.90	6561.0
161.10	8330.0
162.00	2253.0
163.00	686.0
163.80	897.0
165.00	7532.0
166.10	6153.0
167.00	38256.0
168.00	27640.0
169.00	4250.0
170.10	1091.0
170.90	2023.0
171.90	3973.0
173.00	4593.0
174.00	7638.0
175.00	15998.0
176.10	4244.0
176.90	7074.0
177.90	2401.0
179.00	29968.0
180.00	23152.0
181.00	9954.0
182.00	2160.0
183.00	942.0
183.90	2579.0
185.00	14730.0
186.00	112472.0
187.00	34264.0
188.00	3533.0
189.00	6895.0
190.10	1180.0
191.00	4076.0
192.00	9384.0
193.00	11052.0
194.00	2132.0
195.00	1136.0
196.00	24864.0
197.10	6156.0
198.00	886720.0
199.00	59952.0
200.00	4393.0
201.40	3677.0
203.00	6114.0
204.00	30608.0
205.00	52136.0
206.00	229760.0
207.00	38976.0
208.00	8181.0
208.90	3211.0
210.20	3947.0
211.00	8635.0
211.90	2300.0
213.10	636.0
213.90	332.0
214.90	2626.0
215.90	4317.0
217.00	56592.0
218.00	7646.0
219.00	667.0
221.00	35792.0
221.90	8214.0
223.00	12873.0
224.00	128320.0
225.00	32968.0
226.00	3829.0
227.00	52184.0
228.00	7528.0
229.00	11046.0
230.00	1350.0
231.00	4567.0
232.00	631.0
233.00	882.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

234.00	2851.0
234.90	4037.0
236.00	2160.0
237.00	5458.0
238.00	854.0
238.90	2139.0
240.00	1470.0
241.00	3132.0
242.00	6319.0
243.00	6922.0
244.00	91392.0
245.00	13452.0
245.90	18424.0
247.00	3727.0
247.90	931.0
249.00	3316.0
249.90	957.0
250.80	694.0
251.90	1004.0
252.90	1998.0
254.20	4693.0
255.00	496576.0
256.00	75280.0
257.00	6530.0
258.00	25792.0
259.00	3810.0
260.00	843.0
260.90	855.0
263.10	387.0
263.90	548.0
265.00	11422.0
266.00	1556.0
266.70	332.0
269.90	532.0
270.90	1314.0
272.10	1452.0
273.00	13694.0
273.90	36064.0
275.00	215232.0
276.00	27344.0
277.00	16480.0
278.00	3785.0
278.90	652.0
281.00	1204.0
281.90	1072.0
283.10	1967.0
284.00	1605.0
285.00	2654.0
286.00	873.0
288.90	863.0
290.00	578.0
291.00	564.0
291.90	714.0
292.90	3313.0
293.90	1563.0
294.90	1559.0
296.00	56640.0
297.00	8647.0
297.90	562.0
301.10	932.0
301.90	1142.0
303.00	7319.0
303.90	1695.0
305.00	439.0
307.90	720.0
309.00	581.0
310.00	843.0
312.80	695.0
314.00	2842.0
314.90	6287.0
316.00	3147.0
316.80	609.0
321.00	1724.0
321.90	1146.0
323.00	16728.0
324.00	3682.0
325.10	473.0
325.90	433.0
326.90	3266.0
328.00	1996.0
332.00	1388.0
333.10	1695.0
334.00	11363.0
335.00	2908.0
336.00	398.0
341.00	2377.0
341.90	396.0
345.90	4139.0
347.00	993.0
351.00	321.0
351.90	5166.0
352.90	3768.0
354.00	4924.0
354.90	973.0
358.80	372.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

364.00	496.0
365.00	21384.0
365.90	2683.0
367.00	342.0
370.00	492.0
371.00	1383.0
372.00	9189.0
373.00	2016.0
383.00	2229.0
384.00	759.0
384.90	328.0
390.00	1685.0
391.00	1135.0
391.80	379.0
401.00	481.0
401.90	3419.0
403.00	4328.0
403.90	1676.0
404.90	616.0
421.00	3451.0
422.00	3867.0
423.00	30552.0
424.00	10083.0
425.00	1364.0
441.00	69336.0
442.00	429184.0
443.00	83112.0
444.00	7658.0
445.00	969.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

m/z	Abundance
35.90	428.0
37.00	1372.0
38.00	3226.0
39.00	20280.0
40.00	3624.0
41.10	1064.0
42.90	620.0
44.00	10667.0
45.00	904.0
46.90	435.0
47.90	313.0
49.00	3156.0
50.00	70320.0
51.00	271424.0
52.00	15172.0
53.00	922.0
55.00	1916.0
56.00	8622.0
57.00	20576.0
57.90	1106.0
58.90	346.0
59.90	454.0
61.00	4360.0
62.00	4287.0
63.00	13558.0
64.10	2288.0
65.00	7900.0
65.90	578.0
67.10	875.0
68.10	7578.0
69.00	380352.0
70.00	2146.0
72.10	334.0
73.00	3509.0
74.00	38768.0
75.00	60512.0
76.10	19944.0
77.00	436032.0
78.00	32272.0
79.00	27904.0
80.00	21224.0
81.00	30304.0
82.00	8137.0
83.00	6475.0
84.00	2074.0
85.00	6033.0
86.00	11025.0
87.00	3745.0
88.00	1469.0
88.90	845.0
91.00	7063.0
92.00	8102.0
93.00	44744.0
94.00	3678.0
95.00	1166.0
95.90	2638.0
96.90	792.0
98.00	36848.0
99.00	29632.0
100.10	3051.0
101.00	18056.0
102.00	985.0
103.00	5694.0
103.90	11743.0
105.00	9913.0
106.00	3414.0
107.00	123208.0
108.00	19736.0
109.10	4233.0
110.00	238976.0
111.00	35288.0
112.00	4172.0
113.00	1609.0
115.10	641.0
116.00	7182.0
117.00	95856.0
118.00	7906.0
118.90	1256.0
120.10	1839.0
121.00	546.0
122.00	8773.0
123.00	12994.0
124.00	5970.0
125.00	5244.0
126.20	3219.0
127.00	436736.0
128.00	35096.0
129.00	162752.0
130.00	13889.0
130.90	2814.0
132.00	1406.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

132.90	932.0
134.00	5164.0
135.00	14543.0
135.90	5584.0
137.00	8051.0
137.90	1593.0
139.10	960.0
139.90	2881.0
140.90	22552.0
141.90	7623.0
143.00	3978.0
144.00	1166.0
145.00	1237.0
146.00	3466.0
147.00	12608.0
148.00	24680.0
148.90	5624.0
150.00	1519.0
151.10	3185.0
151.80	1948.0
153.00	6111.0
154.00	4609.0
155.00	12404.0
156.00	20568.0
157.00	3767.0
157.90	5033.0
159.00	2964.0
159.90	7672.0
161.00	10950.0
162.00	2778.0
162.90	1153.0
164.00	1181.0
165.00	8372.0
166.00	6807.0
167.00	48056.0
168.00	29056.0
168.90	4592.0
170.00	1685.0
170.90	1968.0
172.00	4588.0
173.00	5532.0
174.00	10825.0
175.00	17712.0
176.00	4980.0
177.00	9619.0
178.00	2779.0
179.00	35584.0
180.00	27976.0
181.00	12033.0
182.00	1713.0
183.10	805.0
184.00	3123.0
185.00	18072.0
186.00	140096.0
187.00	43480.0
188.10	4642.0
189.00	8709.0
190.00	2065.0
191.00	4339.0
192.00	12597.0
193.00	14894.0
194.00	2686.0
195.00	1865.0
196.00	31696.0
197.10	7047.0
197.90	1089024.0
199.00	74608.0
200.00	5485.0
201.30	4770.0
203.00	8119.0
204.00	38448.0
205.00	67368.0
206.00	288576.0
207.00	41152.0
208.00	10795.0
209.00	3655.0
210.10	4590.0
211.00	10014.0
213.00	1088.0
214.20	419.0
215.00	3011.0
215.90	6481.0
217.00	74824.0
218.00	9618.0
218.90	870.0
221.00	47368.0
221.80	10276.0
223.00	17280.0
224.00	162752.0
225.00	41080.0
226.00	5076.0
227.00	63632.0
228.00	8430.0
229.00	13639.0
230.00	1666.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

231.00	5533.0
231.90	1001.0
233.10	942.0
234.00	4575.0
234.90	6316.0
236.00	3183.0
237.00	6345.0
237.90	1217.0
239.00	2547.0
239.90	2033.0
241.00	2884.0
242.00	8221.0
243.00	8527.0
244.00	129360.0
245.00	16456.0
246.00	24632.0
247.00	5477.0
248.00	1373.0
249.00	4977.0
250.00	609.0
250.90	1181.0
252.00	1769.0
253.00	2967.0
255.00	663040.0
256.00	98912.0
257.00	8038.0
258.00	35400.0
259.00	5763.0
259.90	939.0
260.90	1262.0
264.10	908.0
265.00	14481.0
266.00	1877.0
266.80	446.0
269.80	667.0
270.90	1682.0
271.90	1516.0
273.00	20232.0
274.00	50464.0
275.00	303296.0
276.00	37560.0
277.00	23464.0
278.00	4245.0
279.00	915.0
281.00	1273.0
282.00	971.0
283.00	2480.0
284.00	1644.0
285.00	3926.0
286.10	898.0
289.10	983.0
289.90	862.0
290.90	589.0
292.00	1101.0
293.00	5799.0
294.00	1576.0
295.00	1729.0
296.00	77888.0
297.00	10251.0
297.90	719.0
300.90	974.0
302.00	1532.0
303.00	8261.0
304.10	3011.0
305.00	503.0
308.00	1269.0
309.00	898.0
310.00	1111.0
311.00	317.0
312.90	732.0
314.00	3207.0
315.00	9110.0
316.00	4888.0
317.00	1078.0
319.90	490.0
320.90	2606.0
322.10	1293.0
323.00	25088.0
324.00	5078.0
324.90	389.0
325.90	484.0
326.90	5286.0
328.10	2405.0
328.90	476.0
332.10	1476.0
333.00	2519.0
334.00	16080.0
335.00	4055.0
335.90	347.0
339.80	355.0
340.90	2813.0
342.00	798.0
345.90	5733.0
347.00	1247.0
350.80	331.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

352.00	7822.0
353.00	5079.0
354.00	8250.0
355.00	1615.0
359.00	541.0
365.00	32672.0
365.90	4677.0
370.00	679.0
370.80	1849.0
372.00	12744.0
373.00	3220.0
383.00	3803.0
383.90	1096.0
385.10	306.0
390.00	1594.0
390.90	1725.0
392.00	820.0
393.00	319.0
401.00	842.0
402.00	5159.0
403.00	7486.0
404.00	3016.0
404.80	409.0
414.90	481.0
421.00	5942.0
421.90	5907.0
423.00	44816.0
424.00	12981.0
425.00	1566.0
441.00	110456.0
442.00	731968.0
443.00	143936.0
444.00	14278.0
445.00	886.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

m/z	Abundance
37.00	650.0
38.00	1949.0
39.00	10746.0
40.00	2803.0
41.00	1267.0
42.90	547.0
44.00	12378.0
45.10	470.0
46.90	384.0
49.00	2041.0
50.00	39464.0
51.00	157312.0
52.00	7411.0
53.00	424.0
55.00	1161.0
56.00	5340.0
57.00	12023.0
57.90	582.0
59.10	311.0
61.00	2642.0
62.00	2495.0
63.00	8492.0
64.00	1019.0
65.00	3787.0
66.90	569.0
68.00	4054.0
69.00	224768.0
70.00	1616.0
73.00	2686.0
74.00	22768.0
75.00	35200.0
76.00	12656.0
77.00	257152.0
78.00	16323.0
79.00	17336.0
80.00	13465.0
81.00	19624.0
82.00	4423.0
83.00	3986.0
84.00	1569.0
84.90	3649.0
86.00	5694.0
87.00	2434.0
87.90	1022.0
88.80	527.0
91.00	4127.0
92.00	4435.0
93.00	27928.0
93.90	1981.0
94.90	497.0
95.90	1459.0
97.00	1114.0
98.00	20240.0
99.00	19560.0
100.00	2341.0
101.00	9101.0
101.90	649.0
103.00	3367.0
104.00	6267.0
105.00	6532.0
106.00	2157.0
107.00	77296.0
108.00	11625.0
109.10	2619.0
110.00	142528.0
111.00	19624.0
112.00	2241.0
113.00	926.0
114.70	457.0
116.10	3970.0
117.00	60312.0
118.00	4579.0
119.00	769.0
120.00	1180.0
121.10	305.0
122.00	4418.0
122.90	7712.0
124.00	3247.0
125.00	3131.0
126.20	1814.0
127.00	264896.0
128.00	18128.0
129.00	99856.0
130.00	8668.0
131.00	1438.0
132.00	736.0
132.80	1182.0
134.00	3328.0
135.00	8651.0
136.00	3622.0
137.00	4135.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

137.80	1186.0
138.90	545.0
140.00	1435.0
141.00	14523.0
141.90	5756.0
143.00	3332.0
144.00	1213.0
144.80	608.0
145.90	2844.0
147.00	6733.0
148.00	17224.0
149.00	3182.0
150.10	1006.0
151.10	1738.0
151.90	1387.0
153.00	3774.0
154.00	3984.0
155.00	9074.0
156.00	13137.0
157.00	2675.0
157.90	2929.0
158.90	1924.0
160.00	4260.0
161.00	6017.0
162.00	1997.0
162.90	880.0
163.90	1187.0
165.00	5517.0
166.10	4247.0
167.00	29736.0
168.00	19488.0
169.00	2740.0
169.90	791.0
171.00	1311.0
172.00	2545.0
173.00	2991.0
174.00	6109.0
175.00	12433.0
175.90	2853.0
177.00	5777.0
177.90	2100.0
178.90	23376.0
180.00	16744.0
181.00	7470.0
182.00	992.0
182.90	766.0
184.00	2318.0
185.00	11357.0
186.00	95600.0
187.00	25384.0
188.10	2348.0
189.00	5297.0
189.90	1038.0
191.00	3632.0
192.00	7753.0
193.10	8487.0
194.00	1809.0
195.10	1041.0
196.00	19880.0
197.10	5007.0
197.90	714752.0
199.00	48728.0
200.00	3823.0
201.30	2504.0
201.60	2474.0
202.20	806.0
203.00	4961.0
204.00	26136.0
205.00	42048.0
206.00	194304.0
207.00	29040.0
208.00	7056.0
209.00	3210.0
210.00	3107.0
211.00	7136.0
212.20	674.0
212.90	575.0
215.00	2223.0
215.90	4002.0
217.00	51912.0
218.00	7658.0
219.10	684.0
221.00	34560.0
223.00	10829.0
224.00	106736.0
225.00	28264.0
226.00	3544.0
227.00	44016.0
227.90	6214.0
229.00	8593.0
230.10	1554.0
231.00	4335.0
231.90	809.0
232.90	899.0
234.00	2387.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

235.00	3670.0
236.00	2690.0
237.10	4348.0
238.00	550.0
239.00	2149.0
239.90	1393.0
241.00	2516.0
242.00	5944.0
243.10	6644.0
244.00	84792.0
245.00	11660.0
246.00	17392.0
247.00	3200.0
248.10	590.0
249.00	3119.0
249.90	578.0
251.10	855.0
252.10	884.0
253.00	1943.0
255.00	459392.0
256.00	65192.0
257.00	5108.0
258.00	24968.0
259.00	3824.0
260.10	721.0
260.90	429.0
263.00	470.0
263.90	620.0
265.00	10450.0
265.80	1650.0
266.90	356.0
270.00	671.0
270.90	752.0
271.90	1067.0
273.00	14427.0
274.00	37072.0
275.00	215936.0
276.00	27216.0
277.00	15868.0
278.00	3116.0
279.00	810.0
280.90	714.0
282.00	852.0
283.00	1750.0
283.90	1120.0
285.00	2836.0
285.90	749.0
288.80	632.0
289.90	549.0
290.80	314.0
292.10	1031.0
293.00	3178.0
294.00	1600.0
294.90	1616.0
296.00	55944.0
297.00	8328.0
297.90	512.0
301.00	614.0
302.00	1047.0
303.00	6716.0
304.00	1478.0
307.90	815.0
308.90	372.0
309.90	444.0
313.00	827.0
314.00	2896.0
315.00	5701.0
316.00	4313.0
316.90	481.0
320.00	332.0
320.90	1497.0
322.10	772.0
323.00	20384.0
324.10	3138.0
325.00	392.0
325.90	505.0
327.00	3570.0
328.00	2301.0
328.90	550.0
331.90	1502.0
333.00	2003.0
334.00	11347.0
335.00	3241.0
335.90	560.0
338.90	332.0
341.10	2514.0
342.00	720.0
345.90	3871.0
347.00	629.0
351.00	500.0
351.90	5856.0
353.10	3838.0
354.00	5815.0
355.00	1480.0
358.90	354.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52

364.10	370.0
364.90	26056.0
365.90	4173.0
367.00	302.0
370.00	713.0
370.90	1679.0
372.00	11360.0
373.00	2501.0
376.90	561.0
382.90	3511.0
383.90	923.0
385.10	375.0
389.90	1627.0
391.10	1057.0
392.00	628.0
400.80	459.0
401.10	439.0
402.00	4432.0
403.00	6510.0
404.00	3397.0
405.00	332.0
420.90	5289.0
422.00	5093.0
423.00	37208.0
424.00	10530.0
425.00	1150.0
441.00	102680.0
442.00	645760.0
443.00	126680.0
444.00	11836.0
444.90	565.0

Instrument :
BNA_N
ClientSampleId :
DFTPP52