

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
 Data File : BN016428.D
 Acq On : 23 Sep 2021 18:25
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092421

Quant Time: Sep 24 02:31:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 18:55:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	22547	0.400	ng	0.00
7) Naphthalene-d8	10.445	136	97965	0.400	ng	# 0.00
13) Acenaphthene-d10	14.294	164	49725	0.400	ng	0.00
19) Phenanthrene-d10	17.054	188	94665	0.400	ng	# 0.01
29) Chrysene-d12	21.259	240	89783	0.400	ng	# 0.01
35) Perylene-d12	23.519	264	81735	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	18623	0.368	ng	0.00
5) Phenol-d6	6.849	99	21939	0.367	ng	0.00
8) Nitrobenzene-d5	8.810	82	28247	0.407	ng	0.00
11) 2-Methylnaphthalene-d10	12.034	152	54826	0.378	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	4324	0.358	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	78186	0.374	ng	0.00
27) Fluoranthene-d10	19.093	212	93666	0.361	ng	0.00
31) Terphenyl-d14	19.703	244	81169	0.389	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	3680	0.381	ng	# 81
3) n-Nitrosodimethylamine	3.631	42	9216	0.397	ng	# 71
6) bis(2-Chloroethyl)ether	7.107	93	23988	0.397	ng	92
9) Naphthalene	10.483	128	98349	0.376	ng	99
10) Hexachlorobutadiene	10.774	225	20093	0.387	ng	# 99
12) 2-Methylnaphthalene	12.105	142	63785	0.385	ng	97
16) Acenaphthylene	14.015	152	78126	0.386	ng	100
17) Acenaphthene	14.357	154	53265	0.371	ng	98
18) Fluorene	15.347	166	62940	0.365	ng	98
20) 4,6-Dinitro-2-methylph...	15.436	198	2304	0.354	ng	# 66
21) 4-Bromophenyl-phenylether	16.251	248	20245	0.366	ng	97
22) Hexachlorobenzene	16.348	284	20931	0.382	ng	# 90
23) Atrazine	16.531	200	12305	0.370	ng	98
24) Pentachlorophenol	16.701	266	5038	0.423	ng	99
25) Phenanthrene	17.090	178	103556	0.368	ng	99
26) Anthracene	17.176	178	82631	0.352	ng	99
28) Fluoranthene	19.117	202	115842	0.375	ng	99
30) Pyrene	19.485	202	110493	0.393	ng	99
32) Benzo(a)anthracene	21.238	228	100708	0.387	ng	99
33) Chrysene	21.291	228	110139	0.394	ng	99
34) Bis(2-ethylhexyl)phtha...	21.195	149	27368	0.386	ng	97
36) Indeno(1,2,3-cd)pyrene	25.815	276	102185	0.386	ng	99
37) Benzo(b)fluoranthene	22.837	252	97066	0.355	ng	98
38) Benzo(k)fluoranthene	22.882	252	105764	0.401	ng	99
39) Benzo(a)pyrene	23.419	252	81476	0.366	ng	97
40) Dibenzo(a,h)anthracene	25.836	278	88307	0.391	ng	96
41) Benzo(g,h,i)perylene	26.514	276	84876	0.366	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
Data File : BN016428.D
Acq On : 23 Sep 2021 18:25
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN092421

Quant Time: Sep 24 02:31:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 23 18:55:22 2021
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.670 min Scan# 498
 Delta R.T. -0.000 min
 Lab File: BN016428.D
 Acq: 23 Sep 2021 18:25




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092421


Tgt Ion:152 Resp: 22547
 Ion Ratio Lower Upper
 152 100
 150 155.6 125.6 188.4
 115 58.7 50.2 75.2



#2
 1,4-Dioxane
 Concen: 0.381 ng
 RT: 3.281 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN016428.D
 Acq: 23 Sep 2021 18:25

Tgt Ion: 88 Resp: 3680
 Ion Ratio Lower Upper
 88 100
 43 46.4 20.3 30.5#
 58 98.7 69.9 104.9









#3
n-Nitrosodimethylamine
Concen: 0.397 ng
RT: 3.631 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion: 42 Resp: 9216
Ion Ratio Lower Upper
42 100
74 106.7 113.6 170.4#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.368 ng
RT: 5.291 min Scan# 240
Delta R.T. 0.009 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion: 112 Resp: 18623
Ion Ratio Lower Upper
112 100
64 63.2 47.3 70.9
63 32.9 23.4 35.0









#5
Phenol-d6
Concen: 0.367 ng
RT: 6.849 min Scan# 409
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421

Tgt Ion: 99 Resp: 21939
Ion Ratio Lower Upper
99 100
42 25.6 14.0 21.0#
71 31.9 23.8 35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.397 ng
RT: 7.107 min Scan# 437
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion: 93 Resp: 23988
Ion Ratio Lower Upper
93 100
63 88.6 63.4 95.0
95 33.0 26.8 40.2








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.445 min Scan# 736
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421

Tgt Ion:136	Resp:	97965
Ion Ratio	Lower	Upper
136	100	
137	10.9	8.8 13.2
54	5.9	5.3 7.9
68	3.8	3.9 5.9#



#8
Nitrobenzene-d5
Concen: 0.407 ng
RT: 8.810 min Scan# 607
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion: 82	Resp:	28247
Ion Ratio	Lower	Upper
82	100	
128	39.9	31.7 47.5
54	66.8	46.4 69.6









#9
Naphthalene
Concen: 0.376 ng
RT: 10.483 min Scan# 739
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421

Tgt Ion:	128	Resp:	98349
Ion Ratio	Lower	Upper	
128	100		
129	11.0	8.9	13.3
127	13.1	10.1	15.1



#10
Hexachlorobutadiene
Concen: 0.387 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:	225	Resp:	20093
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.5	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.378 ng
RT: 12.034 min Scan# 938
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:152 Resp: 54826
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.385 ng
RT: 12.105 min Scan# 954
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:142 Resp: 63785
Ion Ratio Lower Upper
142 100
141 88.7 69.8 104.8
115 33.9 24.4 36.6








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:164 Resp: 49725
Ion Ratio Lower Upper
164 100
162 101.9 79.0 118.4
160 47.4 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.358 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:330 Resp: 4324
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 42.7 25.0 37.4#









#15
2-Fluorobiphenyl
Concen: 0.374 ng
RT: 12.924 min Scan# 1077
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:	172	Resp:	78186
Ion Ratio	Lower	Upper	
172	100		
171	33.8	27.8	41.6
170	21.8	18.2	27.2



#16
Acenaphthylene
Concen: 0.386 ng
RT: 14.015 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:	152	Resp:	78126
Ion Ratio	Lower	Upper	
152	100		
151	19.2	15.2	22.8
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.371 ng
RT: 14.357 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:154 Resp: 53265
Ion Ratio Lower Upper
154 100
153 112.3 89.3 133.9
152 56.4 42.2 63.4



#18
Fluorene
Concen: 0.365 ng
RT: 15.347 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:166 Resp: 62940
Ion Ratio Lower Upper
166 100
165 97.8 76.7 115.1
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.054 min Scan# 1406
Delta R.T. 0.012 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:188 Resp: 94665
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 5.0 7.0 10.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.354 ng
RT: 15.436 min Scan# 1275
Delta R.T. -0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:198 Resp: 2304
Ion Ratio Lower Upper
198 100
182 13.7 26.6 39.8#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.366 ng
RT: 16.251 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:248	Resp:	20245
Ion Ratio	Lower	Upper
248	100	
250	98.4	80.5 120.7
141	45.8	34.8 52.2



#22
Hexachlorobenzene
Concen: 0.382 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:284	Resp:	20931
Ion Ratio	Lower	Upper
284	100	
142	37.5	22.6 34.0#
249	32.2	24.4 36.6









#23
Atrazine
Concen: 0.370 ng
RT: 16.531 min Scan# 1363
Delta R.T. -0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:	200	Resp:	12305
Ion Ratio	Lower	Upper	
200	100		
173	21.2	18.7	28.1
215	50.9	41.6	62.4



#24
Pentachlorophenol
Concen: 0.423 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:	266	Resp:	5038
Ion Ratio	Lower	Upper	
266	100		
264	63.6	50.6	75.8
268	64.3	52.6	79.0









#25
Phenanthrene
Concen: 0.368 ng
RT: 17.090 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:178 Resp: 103556
Ion Ratio Lower Upper
178 100
176 18.8 14.4 21.6
179 15.5 12.2 18.4



#26
Anthracene
Concen: 0.352 ng
RT: 17.176 min Scan# 1416
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:178 Resp: 82631
Ion Ratio Lower Upper
178 100
176 18.2 13.9 20.9
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.361 ng
RT: 19.093 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:212 Resp: 93666
Ion Ratio Lower Upper
212 100
106 11.1 9.8 14.6
104 6.2 5.4 8.2



#28
Fluoranthene
Concen: 0.375 ng
RT: 19.117 min Scan# 1707
Delta R.T. -0.005 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:202 Resp: 115842
Ion Ratio Lower Upper
202 100
101 9.7 8.5 12.7
203 17.5 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.259 min Scan# 2004
Delta R.T. 0.011 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:240 Resp: 89783
Ion Ratio Lower Upper
240 100
120 4.1 5.3 7.9#
236 25.0 19.8 29.8



#30
Pyrene
Concen: 0.393 ng
RT: 19.485 min Scan# 1781
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:202 Resp: 110493
Ion Ratio Lower Upper
202 100
200 20.5 15.8 23.6
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.389 ng
 RT: 19.703 min Scan# 1825
 Delta R.T. 0.000 min
 Lab File: BN016428.D
 Acq: 23 Sep 2021 18:25




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092421


Tgt Ion:244	Resp:	81169
Ion Ratio	Lower	Upper
244	100	
212	7.3	5.5 8.3
122	9.5	9.0 13.4



#32
 Benzo(a)anthracene
 Concen: 0.387 ng
 RT: 21.238 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: BN016428.D
 Acq: 23 Sep 2021 18:25

Tgt Ion:228	Resp:	100708
Ion Ratio	Lower	Upper
228	100	
226	26.7	20.8 31.2
229	19.4	15.8 23.6









#33
Chrysene
Concen: 0.394 ng
RT: 21.291 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25




Instrument :
BNA_N
ClientSampleId :
ICVBN092421


Tgt Ion:228 Resp: 110139
Ion Ratio Lower Upper
228 100
226 29.0 23.0 34.4
229 19.7 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.386 ng
RT: 21.195 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Tgt Ion:149 Resp: 27368
Ion Ratio Lower Upper
149 100
167 27.4 23.4 35.0
279 4.4 3.9 5.9





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.519 min Scan# 2437

Delta R.T. 0.000 min

Lab File: BN016428.D

Acq: 23 Sep 2021 18:25





Instrument :

BNA_N

ClientSampleId :

ICVBN092421

Tgt Ion:264	Resp:	81735
Ion Ratio	Lower	Upper
264	100	
260	24.5	18.9 28.3
265	24.5	23.7 35.5



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.386 ng




RT: 25.815 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BN016428.D

Acq: 23 Sep 2021 18:25

Tgt Ion:276	Resp:	102185
Ion Ratio	Lower	Upper
276	100	
138	23.3	19.3 28.9
277	25.1	20.2 30.2





#37

Benzo(b)fluoranthene

Concen: 0.355 ng

RT: 22.837 min Scan# 2238

Delta R.T. 0.000 min

Lab File: BN016428.D

Acq: 23 Sep 2021 18:25





Tgt Ion:	252	Resp:	97066
Ion Ratio	Lower	Upper	
252	100		
253	22.4	18.2	27.2
125	9.0	8.4	12.6

Instrument :

BNA_N

ClientSampleId :

ICVBN092421



#38

Benzo(k)fluoranthene

Concen: 0.401 ng




RT: 22.882 min Scan# 2251

Delta R.T. 0.000 min

Lab File: BN016428.D

Acq: 23 Sep 2021 18:25

Tgt Ion:	252	Resp:	105764
Ion Ratio	Lower	Upper	
252	100		
253	22.3	18.2	27.2
125	10.5	8.2	12.4





#39

Benzo(a)pyrene

Concen: 0.366 ng

RT: 23.419 min Scan# 2408

Delta R.T. 0.000 min

Lab File: BN016428.D

Acq: 23 Sep 2021 18:25





Tgt Ion:	252	Resp:	81476
Ion Ratio	Lower	Upper	
252	100		
253	22.5	18.7	28.1
125	10.1	9.9	14.9

Instrument :

BNA_N

ClientSampleId :

ICVBN092421



#40

Dibenzo(a,h)anthracene

Concen: 0.391 ng




RT: 25.836 min Scan# 3114


Delta R.T. 0.000 min

Lab File: BN016428.D

Acq: 23 Sep 2021 18:25

Tgt Ion:	278	Resp:	88307
Ion Ratio	Lower	Upper	
278	100		
139	16.7	14.2	21.2
279	23.7	21.3	31.9





#41
Benzo(g,h,i)perylene
Concen: 0.366 ng
RT: 26.514 min Scan# 3312
Delta R.T. 0.000 min
Lab File: BN016428.D
Acq: 23 Sep 2021 18:25

Instrument :
BNA_N
ClientSampleId :
ICVBN092421

Tgt Ion:	276	Resp:	84876
Ion Ratio	Lower	Upper	
276	100		
277	23.8	20.5	30.7
138	19.8	18.2	27.4

