

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
 Data File : BN016429.D
 Acq On : 23 Sep 2021 19:39
 Operator : CG/JU
 Sample : PB139285BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB139285BL

Quant Time: Sep 24 02:32:31 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 18:55:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	23062	0.400	ng	0.00
7) Naphthalene-d8	10.432	136	101060	0.400	ng	#-0.01
13) Acenaphthene-d10	14.294	164	49912	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	90358	0.400	ng	# 0.00
29) Chrysene-d12	21.259	240	72446	0.400	ng	# 0.01
35) Perylene-d12	23.519	264	62666	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.310	112	16802	0.325	ng	0.03
5) Phenol-d6	6.859	99	18040	0.295	ng	0.00
8) Nitrobenzene-d5	8.810	82	24064	0.336	ng	0.00
11) 2-Methylnaphthalene-d10	12.034	152	50065	0.334	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	2405	0.253	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	68635	0.327	ng	0.00
27) Fluoranthene-d10	19.088	212	81729	0.330	ng	0.00
31) Terphenyl-d14	19.703	244	68452	0.406	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
Data File : BN016429.D
Acq On : 23 Sep 2021 19:39
Operator : CG/JU
Sample : PB139285BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB139285BL

Quant Time: Sep 24 02:32:31 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 23 18:55:22 2021
Response via : Initial Calibration

