

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
 Data File : BN016446.D
 Acq On : 24 Sep 2021 20:24
 Operator : CG/JU
 Sample : M3911-04
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SW-CB-02-(0.2)

Quant Time: Sep 25 05:06:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	22529	0.40	ng	0.00
7) Naphthalene-d8	10.432	136	99892	0.40	ng	#-0.01
13) Acenaphthene-d10	14.294	164	53320	0.40	ng	0.00
19) Phenanthrene-d10	17.042	188	109813	0.40	ng	0.00
29) Chrysene-d12	21.248	240	105607	0.40	ng	# 0.00
35) Perylene-d12	23.518	264	95479	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	7530	0.15	ng	0.02
5) Phenol-d6	6.849	99	5189	0.09	ng	0.00
8) Nitrobenzene-d5	8.809	82	21800	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.029	152	47739	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	6230	0.44	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	62153	0.28	ng	0.00
27) Fluoranthene-d10	19.087	212	126373	0.42	ng	0.00
31) Terphenyl-d14	19.698	244	101085	0.41	ng	0.00
Target Compounds						
23) Atrazine	16.530	200	266	0.14	ng	# 37
24) Pentachlorophenol	16.701	266	269	0.25	ng	87
32) Benzo(a)anthracene	21.238	228	19241	0.06	ng	96
33) Chrysene	21.291	228	19320	0.06	ng	97
34) Bis(2-ethylhexyl)phtha...	21.185	149	40449	0.48	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
Data File : BN016446.D
Acq On : 24 Sep 2021 20:24
Operator : CG/JU
Sample : M3911-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SW-CB-02-(0.2)

Quant Time: Sep 25 05:06:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 03:35:41 2021
Response via : Initial Calibration

