

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
 Data File : BN016456.D
 Acq On : 25 Sep 2021 03:38
 Operator : CG/JU
 Sample : M3782-11
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BPOW1-7-20210914

Quant Time: Sep 25 05:07:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	20294	0.40	ng	0.00
7) Naphthalene-d8	10.432	136	90131	0.40	ng	#-0.01
13) Acenaphthene-d10	14.294	164	49002	0.40	ng	0.00
19) Phenanthrene-d10	17.042	188	97110	0.40	ng	0.00
29) Chrysene-d12	21.249	240	94154	0.40	ng	# 0.00
35) Perylene-d12	23.512	264	82872	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	6706	0.15	ng	0.02
5) Phenol-d6	6.849	99	4530	0.08	ng	0.00
8) Nitrobenzene-d5	8.810	82	21400	0.34	ng	0.00
11) 2-Methylnaphthalene-d10	12.029	152	44369	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	4984	0.40	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	59512	0.29	ng	-0.01
27) Fluoranthene-d10	19.088	212	119713	0.45	ng	0.00
31) Terphenyl-d14	19.698	244	99426	0.45	ng	0.00
Target Compounds						
24) Pentachlorophenol	16.701	266	57	0.25	ng	# 61
34) Bis(2-ethylhexyl)phtha...	21.185	149	41863	0.56	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
Data File : BN016456.D
Acq On : 25 Sep 2021 03:38
Operator : CG/JU
Sample : M3782-11
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW1-7-20210914

Quant Time: Sep 25 05:07:20 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 03:35:41 2021
Response via : Initial Calibration

