


2,4,6-Tribromophenol

Response Ratio



$$\text{Response} = 1.480\text{e-}001 * \text{Amt} - 4.557\text{e-}002$$


Coef of Det (r^2) = 0.997003 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN092421.M

Calibration Table Last Updated: Fri Sep 24 03:35:41 2021

4,6-Dinitro-2-methylphenol


Response Ratio



$R = 2.418e-003 A^2 + 4.401e-002 A - 1.654e-002$
 Coef of Det (r^2) = 0.999451 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN092421.M
 Calibration Table Last Updated: Fri Sep 24 03:35:41 2021

Atrazine

Response Ratio



$$\text{Response} = 2.213\text{e-}001 * \text{Amt} - 7.475\text{e-}002$$


Coef of Det (r^2) = 0.998846 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN092421.M

Calibration Table Last Updated: Fri Sep 24 03:35:41 2021

Pentachlorophenol

Response Ratio



$$\text{Response} = 1.190\text{e-}001 * \text{Amt} - 7.269\text{e-}002$$

Coef of Det (r^2) = 0.996759 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN092421.M

Calibration Table Last Updated: Fri Sep 24 03:35:41 2021