

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
 Data File : BN016421.D
 Acq On : 23 Sep 2021 14:08
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 9/27/2021 2:59:47 PM

Quant Time: Sep 23 16:04:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 16:03:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.670	152	20200	0.400	ng	0.00
7) Naphthalene-d8	10.445	136	85064	0.400	ng	# 0.00
13) Acenaphthene-d10	14.294	164	41811	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	79434	0.400	ng	# 0.00
29) Chrysene-d12	21.259	240	70217	0.400	ng	# 0.01
35) Perylene-d12	23.522	264	63011	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.282	112	4297	0.085	ng	0.00
5) Phenol-d6	6.849	99	4887	0.080	ng	0.00
8) Nitrobenzene-d5	8.810	82	5367	0.106	ng	0.00
11) 2-Methylnaphthalene-d10	12.034	152	12098	0.090	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	884	0.072	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	18589	0.114	ng	0.00
27) Fluoranthene-d10	19.088	212	19398	0.084	ng	0.00
31) Terphenyl-d14	19.703	244	18353	0.113	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.281	88	742m	0.088	ng	
3) n-Nitrosodimethylamine	3.631	42	1653	0.086	ng	# 79
6) bis(2-Chloroethyl)ether	7.108	93	5361	0.098	ng	92
9) Naphthalene	10.483	128	22764	0.099	ng	98
10) Hexachlorobutadiene	10.774	225	4461	0.102	ng	# 99
12) 2-Methylnaphthalene	12.109	142	14107	0.093	ng	98
16) Acenaphthylene	14.015	152	13673	0.070	ng	99
17) Acenaphthene	14.358	154	11456	0.095	ng	97
18) Fluorene	15.347	166	13401	0.090	ng	98
20) 4,6-Dinitro-2-methylph...	15.436	198	365	0.110	ng	90
21) 4-Bromophenyl-phenylether	16.251	248	4162	0.090	ng	96
22) Hexachlorobenzene	16.348	284	4462	0.102	ng	# 90
23) Atrazine	16.531	200	2530	0.059	ng	99
25) Phenanthrene	17.091	178	23281m	0.104	ng	
26) Anthracene	17.176	178	16646	0.077	ng	99
28) Fluoranthene	19.118	202	23073	0.089	ng	99
30) Pyrene	19.485	202	23236	0.105	ng	99
32) Benzo(a)anthracene	21.238	228	19665	0.087	ng	99
33) Chrysene	21.291	228	23439	0.107	ng	99
34) Bis(2-ethylhexyl)phtha...	21.196	149	6049	0.050	ng	98
36) Indeno(1,2,3-cd)pyrene	25.822	276	19702	0.097	ng	100
37) Benzo(b)fluoranthene	22.838	252	21059	0.105	ng	97
38) Benzo(k)fluoranthene	22.885	252	20497	0.106	ng	# 94
39) Benzo(a)pyrene	23.423	252	15970	0.088	ng	95
40) Dibenzo(a,h)anthracene	25.839	278	16995	0.098	ng	98
41) Benzo(g,h,i)perylene	26.521	276	17400	0.101	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
Data File : BN016421.D
Acq On : 23 Sep 2021 14:08
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
9/27/2021 2:59:47 PM


Quant Time: Sep 23 16:04:05 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 23 16:03:09 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.670 min Scan# 498
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion: 152 Resp: 20200
Ion Ratio Lower Upper
152 100
150 155.7 125.6 188.4
115 57.9 50.2 75.2

Manual Integrations
APPROVED



mohammad
9/27/2021 2:59:47 PM




#2
1,4-Dioxane
Concen: 0.088 ng m
RT: 3.281 min Scan# 22
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08



Tgt Ion: 88 Resp: 742
Ion Ratio Lower Upper
88 100
43 118.2 20.3 30.5#
58 112.0 69.9 104.9#









#3
n-Nitrosodimethylamine
Concen: 0.086 ng
RT: 3.631 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion: 42 Resp: 1653
Ion Ratio Lower Upper
42 100
74 116.4 113.6 170.4
44 11.5 0.0 0.0#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM



#4
2-Fluorophenol
Concen: 0.085 ng
RT: 5.282 min Scan# 239
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08


Tgt Ion: 112 Resp: 4297
Ion Ratio Lower Upper
112 100
64 62.8 47.3 70.9
63 34.0 23.4 35.0





#5
Phenol-d6
Concen: 0.080 ng
RT: 6.849 min Scan# 409
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Instrument :
BNA_N
ClientSampled :
SSTDICC0.1




Tgt Ion: 99 Resp: 4887
Ion Ratio Lower Upper
99 100
42 24.6 14.0 21.0#
71 32.4 23.8 35.6

Manual Integrations
APPROVED



mohammad
9/27/2021 2:59:47 PM




#6
bis(2-Chloroethyl)ether
Concen: 0.098 ng
RT: 7.108 min Scan# 437
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion: 93 Resp: 5361
Ion Ratio Lower Upper
93 100
63 88.5 63.4 95.0
95 34.2 26.8 40.2





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.445 min Scan# 736
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	136	Resp:	85064
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	5.8	5.3	7.9
68	3.8	3.9	5.9#




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM




#8
Nitrobenzene-d5
Concen: 0.106 ng
RT: 8.810 min Scan# 607
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:	82	Resp:	5367
Ion	Ratio	Lower	Upper
82	100		
128	38.0	31.7	47.5
54	67.9	46.4	69.6








#9
Naphthalene
Concen: 0.099 ng
RT: 10.483 min Scan# 739
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion:128 Resp: 22764

Ion	Ratio	Lower	Upper
128	100		
129	11.5	8.9	13.3
127	13.5	10.1	15.1





#10
Hexachlorobutadiene
Concen: 0.102 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08



Tgt Ion:225 Resp: 4461


Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	51.5	77.3



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
9/27/2021 2:59:47 PM







#11
2-Methylnaphthalene-d10
Concen: 0.090 ng
RT: 12.034 min Scan# 938
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:152 Resp: 12098
Ion Ratio Lower Upper
152 100
151 20.4 16.1 24.1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM




#12
2-Methylnaphthalene
Concen: 0.093 ng
RT: 12.109 min Scan# 955
Delta R.T. 0.005 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:142 Resp: 14107
Ion Ratio Lower Upper
142 100
141 86.6 69.8 104.8
115 32.8 24.4 36.6






#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08



Tgt Ion:164 Resp: 41811


Ion	Ratio	Lower	Upper
164	100		
162	102.2	79.0	118.4
160	46.6	34.1	51.1




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
9/27/2021 2:59:47 PM






#14
2,4,6-Tribromophenol
Concen: 0.072 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08



Tgt Ion:330 Resp: 884

Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	42.8	25.0	37.4









#15
2-Fluorobiphenyl
Concen: 0.114 ng
RT: 12.924 min Scan# 1077
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:	172	Resp:	18589
Ion Ratio	Lower	Upper	
172	100		
171	34.2	27.8	41.6
170	22.4	18.2	27.2

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM



#16
Acenaphthylene
Concen: 0.070 ng
RT: 14.015 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08


Tgt Ion:	152	Resp:	13673
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.2	22.8
153	13.1	10.5	15.7





#17
Acenaphthene
Concen: 0.095 ng
RT: 14.358 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:154 Resp: 11456
Ion Ratio Lower Upper
154 100
153 113.3 89.3 133.9
152 56.1 42.2 63.4




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM




#18
Fluorene
Concen: 0.090 ng
RT: 15.347 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:166 Resp: 13401
Ion Ratio Lower Upper
166 100
165 98.5 76.7 115.1
167 13.4 10.9 16.3







#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion:188 Resp: 79434
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.6 7.0 10.6#




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED



mohammad
9/27/2021 2:59:47 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.110 ng
RT: 15.436 min Scan# 1275
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion:198 Resp: 365
Ion Ratio Lower Upper
198 100
182 27.5 26.6 39.8
77 0.0 0.0 0.0





#21
4-Bromophenyl-phenylether
Concen: 0.090 ng
RT: 16.251 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:248 Resp: 4162
Ion Ratio Lower Upper
248 100
250 99.0 80.5 120.7
141 50.1 34.8 52.2

Manual Integrations
APPROVED



mohammad
9/27/2021 2:59:47 PM




#22
Hexachlorobenzene
Concen: 0.102 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion:284 Resp: 4462
Ion Ratio Lower Upper
284 100
142 37.2 22.6 34.0#
249 32.6 24.4 36.6





#23
Atrazine
Concen: 0.059 ng
RT: 16.531 min Scan# 1363
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion:200 Resp: 2530
Ion Ratio Lower Upper
200 100
173 23.3 18.7 28.1
215 50.9 41.6 62.4

Manual Integrations
APPROVED



mohammad
9/27/2021 2:59:47 PM




#25
Phenanthrene
Concen: 0.104 ng m
RT: 17.091 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion:178 Resp: 23281
Ion Ratio Lower Upper
178 100
176 13.4 14.4 21.6#
179 11.4 12.2 18.4#





#26
 Anthracene
 Concen: 0.077 ng
 RT: 17.176 min Scan# 1416
 Delta R.T. 0.000 min
 Lab File: BN016421.D
 Acq: 23 Sep 2021 14:08

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1





Tgt Ion:178 Resp: 16646


Ion	Ratio	Lower	Upper
178	100		
176	18.0	13.9	20.9
179	15.1	12.2	18.4

Manual Integrations
 APPROVED

mohammad
 9/27/2021 2:59:47 PM






#27
 Fluoranthene-d10
 Concen: 0.084 ng
 RT: 19.088 min Scan# 1701
 Delta R.T. -0.005 min
 Lab File: BN016421.D
 Acq: 23 Sep 2021 14:08



Tgt Ion:212 Resp: 19398


Ion	Ratio	Lower	Upper
212	100		
106	11.4	9.8	14.6
104	6.2	5.4	8.2





#28
Fluoranthene
Concen: 0.089 ng
RT: 19.118 min Scan# 1707
Delta R.T. -0.005 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:202 Resp: 23073
Ion Ratio Lower Upper
202 100
101 9.7 8.5 12.7
203 17.3 13.8 20.8

Manual Integrations
APPROVED



mohammad
9/27/2021 2:59:47 PM




#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.259 min Scan# 2004
Delta R.T. 0.011 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion:240 Resp: 70217
Ion Ratio Lower Upper
240 100
120 5.2 5.3 7.9#
236 25.2 19.8 29.8









#30
Pyrene
Concen: 0.105 ng
RT: 19.485 min Scan# 1781
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08





Tgt Ion:202 Resp: 23236
Ion Ratio Lower Upper
202 100
200 20.2 15.8 23.6
203 17.5 13.9 20.9



#31
Terphenyl-d14
Concen: 0.113 ng
RT: 19.703 min Scan# 1825
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion:244 Resp: 18353
Ion Ratio Lower Upper
244 100
212 7.6 5.5 8.3
122 9.5 9.0 13.4



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
9/27/2021 2:59:47 PM







#32
Benzo(a)anthracene
Concen: 0.087 ng
RT: 21.238 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:228 Resp: 19665
Ion Ratio Lower Upper
228 100
226 27.1 20.8 31.2
229 19.8 15.8 23.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM



#33
Chrysene
Concen: 0.107 ng
RT: 21.291 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:228 Resp: 23439
Ion Ratio Lower Upper
228 100
226 29.7 23.0 34.4
229 19.3 15.6 23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.050 ng
RT: 21.196 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:149 Resp: 6049
Ion Ratio Lower Upper
149 100
167 27.9 23.4 35.0
279 4.4 3.9 5.9

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.522 min Scan# 2438
Delta R.T. 0.004 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:264 Resp: 63011
Ion Ratio Lower Upper
264 100
260 24.4 18.9 28.3
265 26.2 23.7 35.5









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.097 ng
RT: 25.822 min Scan# 3110
Delta R.T. 0.007 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:	276	Resp:	19702
Ion Ratio	Lower	Upper	
276	100		
138	23.8	19.3	28.9
277	25.1	20.2	30.2

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
9/27/2021 2:59:47 PM




#37
Benzo(b)fluoranthene
Concen: 0.105 ng
RT: 22.838 min Scan# 2238
Delta R.T. 0.000 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion:	252	Resp:	21059
Ion Ratio	Lower	Upper	
252	100		
253	24.3	18.2	27.2
125	11.1	8.4	12.6









#38
Benzo(k)fluoranthene
Concen: 0.106 ng
RT: 22.885 min Scan# 2252
Delta R.T. 0.004 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08





Tgt Ion: 252 Resp: 20497
Ion Ratio Lower Upper
252 100
253 24.7 18.2 27.2
125 13.6 8.2 12.4#



#39
Benzo(a)pyrene
Concen: 0.088 ng
RT: 23.423 min Scan# 2409
Delta R.T. 0.004 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08




Tgt Ion: 252 Resp: 15970
Ion Ratio Lower Upper
252 100
253 26.0 18.7 28.1
125 13.6 9.9 14.9



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
9/27/2021 2:59:47 PM







#40
Dibenzo(a,h)anthracene
Concen: 0.098 ng
RT: 25.839 min Scan# 3115
Delta R.T. 0.004 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion: 278 Resp: 16995
Ion Ratio Lower Upper
278 100
139 18.8 14.2 21.2
279 25.9 21.3 31.9

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
9/27/2021 2:59:47 PM



#41
Benzo(g,h,i)perylene
Concen: 0.101 ng
RT: 26.521 min Scan# 3314
Delta R.T. 0.007 min
Lab File: BN016421.D
Acq: 23 Sep 2021 14:08

Tgt Ion: 276 Resp: 17400
Ion Ratio Lower Upper
276 100
277 24.9 20.5 30.7
138 20.7 18.2 27.4

