

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
 Data File : BN016435.D
 Acq On : 24 Sep 2021 11:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Sep 24 11:46:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 03:35:41 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.661	152	21424	0.40	ng	0.00
7) Naphthalene-d8	10.432	136	91928	0.40	ng	#-0.01
13) Acenaphthene-d10	14.294	164	46355	0.40	ng	0.00
19) Phenanthrene-d10	17.042	188	88299	0.40	ng	0.00
29) Chrysene-d12	21.259	240	82783	0.40	ng	# 0.01
35) Perylene-d12	23.522	264	70628	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	16513	0.34	ng	0.00
5) Phenol-d6	6.840	99	19154	0.34	ng	0.00
8) Nitrobenzene-d5	8.810	82	26033	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.030	152	51136	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	3679	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	72111	0.37	ng	-0.01
27) Fluoranthene-d10	19.088	212	87314	0.36	ng	0.00
31) Terphenyl-d14	19.703	244	74396	0.39	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.281	88	3756	0.41	ng	# 94
3) n-Nitrosodimethylamine	3.622	42	9120	0.41	ng	# 69
6) bis(2-Chloroethyl)ether	7.108	93	23079	0.40	ng	93
9) Naphthalene	10.483	128	91534	0.37	ng	99
10) Hexachlorobutadiene	10.774	225	18895	0.39	ng	# 99
12) 2-Methylnaphthalene	12.101	142	59358	0.38	ng	98
16) Acenaphthylene	14.015	152	71059	0.38	ng	99
17) Acenaphthene	14.358	154	49611	0.37	ng	98
18) Fluorene	15.347	166	59016	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.436	198	2231	0.36	ng	# 70
21) 4-Bromophenyl-phenylether	16.251	248	18783	0.36	ng	97
22) Hexachlorobenzene	16.348	284	19464	0.38	ng	# 89
23) Atrazine	16.519	200	10434	0.35	ng	# 91
24) Pentachlorophenol	16.701	266	4498	0.42	ng	99
25) Phenanthrene	17.091	178	97011	0.37	ng	99
26) Anthracene	17.176	178	75735	0.35	ng	99
28) Fluoranthene	19.118	202	108021	0.38	ng	99
30) Pyrene	19.480	202	102091	0.39	ng	99
32) Benzo(a)anthracene	21.238	228	89761	0.37	ng	99
33) Chrysene	21.291	228	102273	0.40	ng	100
34) Bis(2-ethylhexyl)phtha...	21.196	149	24901	0.38	ng	97
36) Indeno(1,2,3-cd)pyrene	25.819	276	82731	0.36	ng	99
37) Benzo(b)fluoranthene	22.838	252	86860	0.37	ng	98
38) Benzo(k)fluoranthene	22.882	252	90261	0.40	ng	99
39) Benzo(a)pyrene	23.420	252	69518	0.36	ng	97
40) Dibenzo(a,h)anthracene	25.840	278	70710	0.36	ng	95
41) Benzo(g,h,i)perylene	26.514	276	71182	0.36	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092421\
Data File : BN016435.D
Acq On : 24 Sep 2021 11:06
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Sep 24 11:46:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 03:35:41 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.661 min Scan# 497
Delta R.T. -0.009 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 21424
Ion Ratio Lower Upper
152 100
150 157.9 125.6 188.4
115 60.0 50.2 75.2



#2
1,4-Dioxane
Concen: 0.41 ng
RT: 3.281 min Scan# 22
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion: 88 Resp: 3756
Ion Ratio Lower Upper
88 100
43 37.5 20.3 30.5#
58 88.4 69.9 104.9









#3
n-Nitrosodimethylamine
Concen: 0.41 ng
RT: 3.622 min Scan# 59
Delta R.T. -0.009 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 9120
Ion Ratio Lower Upper
42 100
74 103.7 113.6 170.4#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.34 ng
RT: 5.291 min Scan# 240
Delta R.T. 0.010 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion: 112 Resp: 16513
Ion Ratio Lower Upper
112 100
64 63.4 47.3 70.9
63 33.6 23.4 35.0









#5
Phenol-d6
Concen: 0.34 ng
RT: 6.840 min Scan# 408
Delta R.T. -0.009 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 19154
Ion Ratio Lower Upper
99 100
42 26.7 14.0 21.0#
71 32.1 23.8 35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.40 ng
RT: 7.108 min Scan# 437
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion: 93 Resp: 23079
Ion Ratio Lower Upper
93 100
63 88.2 63.4 95.0
95 33.2 26.8 40.2









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.432 min Scan# 735
Delta R.T. -0.012 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	91928
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	9.3	5.3	7.9#
68	5.6	3.9	5.9



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.810 min Scan# 607
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:	82	Resp:	26033
Ion	Ratio	Lower	Upper
82	100		
128	45.9	31.7	47.5
54	60.5	46.4	69.6









#9
Naphthalene
Concen: 0.37 ng
RT: 10.483 min Scan# 739
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	91534
Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.9	13.3
127	13.0	10.1	15.1



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:	225	Resp:	18895
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 12.030 min Scan# 937
Delta R.T. -0.004 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 51136
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.101 min Scan# 953
Delta R.T. -0.004 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:142 Resp: 59358
Ion Ratio Lower Upper
142 100
141 88.0 69.8 104.8
115 33.8 24.4 36.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.294 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 46355
Ion Ratio Lower Upper
164 100
162 101.1 79.0 118.4
160 46.6 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:330 Resp: 3679
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 43.4 25.0 37.4#









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.012 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	72111
Ion Ratio	Lower	Upper
172	100	
171	35.5	27.8 41.6
170	24.5	18.2 27.2



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.015 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:152	Resp:	71059
Ion Ratio	Lower	Upper
152	100	
151	19.4	15.2 22.8
153	13.0	10.5 15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.358 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 49611
Ion Ratio Lower Upper
154 100
153 111.5 89.3 133.9
152 56.5 42.2 63.4



#18
Fluorene
Concen: 0.37 ng
RT: 15.347 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:166 Resp: 59016
Ion Ratio Lower Upper
166 100
165 97.2 76.7 115.1
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.042 min Scan# 1405
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 88299
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.1 7.0 10.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.36 ng
RT: 15.436 min Scan# 1275
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:198 Resp: 2231
Ion Ratio Lower Upper
198 100
182 16.3 26.6 39.8#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.251 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:248 Resp: 18783




Ion	Ratio	Lower	Upper
248	100		
250	101.1	80.5	120.7
141	38.8	34.8	52.2




#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:284 Resp: 19464

Ion	Ratio	Lower	Upper
284	100		
142	38.6	22.6	34.0#
249	32.6	24.4	36.6









#23
Atrazine
Concen: 0.35 ng
RT: 16.519 min Scan# 1362
Delta R.T. -0.012 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 10434
Ion Ratio Lower Upper
200 100
173 30.7 18.7 28.1#
215 47.5 41.6 62.4



#24
Pentachlorophenol
Concen: 0.42 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:266 Resp: 4498
Ion Ratio Lower Upper
266 100
264 63.3 50.6 75.8
268 64.5 52.6 79.0









#25
Phenanthrene
Concen: 0.37 ng
RT: 17.091 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 97011
Ion Ratio Lower Upper
178 100
176 18.7 14.4 21.6
179 15.6 12.2 18.4



#26
Anthracene
Concen: 0.35 ng
RT: 17.176 min Scan# 1416
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:178 Resp: 75735
Ion Ratio Lower Upper
178 100
176 18.3 13.9 20.9
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.088 min Scan# 1701
Delta R.T. -0.005 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 87314
Ion Ratio Lower Upper
212 100
106 11.2 9.8 14.6
104 6.2 5.4 8.2



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.118 min Scan# 1707
Delta R.T. -0.005 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:202 Resp: 108021
Ion Ratio Lower Upper
202 100
101 9.8 8.5 12.7
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.259 min Scan# 2004
Delta R.T. 0.011 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 82783
Ion Ratio Lower Upper
240 100
120 4.6 5.3 7.9#
236 25.4 19.8 29.8



#30
Pyrene
Concen: 0.39 ng
RT: 19.480 min Scan# 1780
Delta R.T. -0.005 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:202 Resp: 102091
Ion Ratio Lower Upper
202 100
200 20.4 15.8 23.6
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.39 ng
RT: 19.703 min Scan# 1825
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 74396
Ion Ratio Lower Upper
244 100
212 7.3 5.5 8.3
122 9.0 9.0 13.4



#32
Benzo(a)anthracene
Concen: 0.37 ng
RT: 21.238 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:228 Resp: 89761
Ion Ratio Lower Upper
228 100
226 26.8 20.8 31.2
229 19.6 15.8 23.6









#33
Chrysene
Concen: 0.40 ng
RT: 21.291 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:228 Resp: 102273




Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.0	34.4
229	19.5	15.6	23.4




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.38 ng
RT: 21.196 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:149 Resp: 24901


Ion	Ratio	Lower	Upper
149	100		
167	27.6	23.4	35.0
279	4.1	3.9	5.9








#35
Perylene-d12
Concen: 0.40 ng
RT: 23.522 min Scan# 2438
Delta R.T. 0.004 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:264 Resp: 70628
Ion Ratio Lower Upper
264 100
260 24.7 18.9 28.3
265 24.9 23.7 35.5



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.36 ng
RT: 25.819 min Scan# 3109
Delta R.T. 0.004 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:276 Resp: 82731
Ion Ratio Lower Upper
276 100
138 23.4 19.3 28.9
277 25.0 20.2 30.2









#37
Benzo(b)fluoranthene
Concen: 0.37 ng
RT: 22.838 min Scan# 2238
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Tgt Ion:252 Resp: 86860
Ion Ratio Lower Upper
252 100
253 22.1 18.2 27.2
125 9.2 8.4 12.6


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



#38
Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 22.882 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:252 Resp: 90261
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.2
125 10.4 8.2 12.4









#39
Benzo(a)pyrene
Concen: 0.36 ng
RT: 23.420 min Scan# 2408
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 69518
Ion Ratio Lower Upper
252 100
253 22.6 18.7 28.1
125 10.3 9.9 14.9



#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.840 min Scan# 3115
Delta R.T. 0.004 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06


Tgt Ion:278 Resp: 70710
Ion Ratio Lower Upper
278 100
139 15.7 14.2 21.2
279 23.9 21.3 31.9





#41
Benzo(g,h,i)perylene
Concen: 0.36 ng
RT: 26.514 min Scan# 3312
Delta R.T. 0.000 min
Lab File: BN016435.D
Acq: 24 Sep 2021 11:06

Tgt Ion:	276	Resp:	71182
Ion	Ratio	Lower	Upper
276	100		
277	23.8	20.5	30.7
138	19.7	18.2	27.4



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4